

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

# Humanitarian Device Exemption (HDE) Program

---

## Draft Guidance for Industry and Food and Drug Administration Staff

*DRAFT GUIDANCE*

**This draft guidance document is being distributed for comment purposes only.**

**Document issued on June 13, 2018.**

You should submit comments and suggestions regarding this draft document within 60 days of publication in the *Federal Register* of the notice announcing the availability of the draft guidance. Submit electronic comments to <https://www.regulations.gov>. Submit written comments to the Dockets Management Staff (HFA-305), Food and Drug Administration, 5630 Fishers Lane, rm. 1061, Rockville, MD 20852. Identify all comments with the docket number listed in the notice of availability that publishes in the *Federal Register*.

For questions about this document regarding CDRH-regulated devices, contact HDE Staff, Center for Devices and Radiological Health, at 301-796-5640. For questions about this document regarding CBER-regulated devices, contact the Office of Communication, Outreach, and Development (OCOD) at 1-800-835-4709 or 240-402-8010.

**When final, this guidance will supersede “Guidance for HDE holders, Institutional Review Boards (IRBs), Clinical Investigators, and Food and Drug Administration Staff, Humanitarian Device Exemptions (HDE) Regulation: Questions and Answers,” issued July 8, 2010.**

## Preface

23

### 24 **Additional Copies**

#### 25 **CDRH**

26 Additional copies are available from the Internet. You may also send an email request to [CDRH-](mailto:CDRH-Guidance@fda.hhs.gov)  
27 [Guidance@fda.hhs.gov](mailto:CDRH-Guidance@fda.hhs.gov) to receive a copy of the guidance. Please use the document number  
28 17040 to identify the guidance you are requesting.

#### 29 **CBER**

30 Additional copies are available from the Center for Biologics Evaluation and Research (CBER),  
31 by written request, Office of Communication, Outreach, and Development (OCOD), 10903 New  
32 Hampshire Ave., Bldg. 71, Room 3128, Silver Spring, MD 20993-0002, or by calling 1-800-835-  
33 4709 or 240-402-8010, by email, [ocod@fda.hhs.gov](mailto:ocod@fda.hhs.gov) or from the Internet at [https://www.fda.gov/  
34 BiologicsBloodVaccines/GuidanceComplianceRegulatoryInformation/Guidances/default.htm](https://www.fda.gov/BiologicsBloodVaccines/GuidanceComplianceRegulatoryInformation/Guidances/default.htm).

#### 35 **OOPD**

36 Additional copies of this guidance document are also available from the Office of Orphan  
37 Products Development (OOPD), Office of Special Medical Programs, Food and Drug  
38 Administration, 10903 New Hampshire Ave, Silver Spring, MD 20993, or by calling 301-796-  
39 8660, or from the Internet at [https://www.fda.gov/ForIndustry/DevelopingProductsforRare  
40 DiseasesConditions/DesignatingHumanitarianUseDevicesHUDS/default.htm](https://www.fda.gov/ForIndustry/DevelopingProductsforRareDiseasesConditions/DesignatingHumanitarianUseDevicesHUDS/default.htm). You may also send  
41 an e-mail request to [orphan@fda.hhs.gov](mailto:orphan@fda.hhs.gov).

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

**Table of Contents**

42

43 I. Introduction..... 1

44 II. Background..... 2

45 III. Scope..... 4

46 IV. HUD Designations and HDE Applications..... 5

47 V. FDA’s Review of HDE Applications ..... 5

48 A. HDE Application Required Elements and Filing Review Principles ..... 8

49 B. FDA Review Actions for an HDE Application ..... 11

50 VI. Assessing Probable Benefit and Risk in an HDE Application ..... 14

51 VII. Post-Approval Requirements..... 17

52 A. IRB or Appropriate Local Committee Oversight and Approval..... 18

53 B. Adverse Event Reporting..... 18

54 C. HDE Supplements..... 19

55 D. HDE Periodic Reports..... 20

56 E. HUD Designation Re-Evaluation ..... 21

57 VIII. Special Considerations for Devices Marketed Under an HDE..... 21

58 A. Eligibility for Profit..... 21

59 B. The Annual Distribution Number (ADN)..... 23

60 C. Information to Patients..... 25

61 D. HDEs and Pediatric Patients ..... 26

62 E. Review and Approval of the Use of HUDs in Clinical Care ..... 27

63 F. Review and Approval for Clinical Testing of HUDs..... 30

64 G. Emergency Use of HUDs..... 31

65 Appendix A – Checklist for Filing Review for HDEs..... 33

66 Appendix B – Considerations for the Probable Benefit-Risk Assessment..... 42

67 Assessment of Probable Benefit ..... 42

68 Assessment of Risk..... 45

69 Assessment of Probable Benefit-Risk ..... 46

70 Appendix C – Probable Benefit-Risk Assessment Tool..... 51

71

# Humanitarian Device Exemption (HDE) Program

## Draft Guidance for Industry and Food and Drug Administration Staff

*This draft guidance, when finalized, will represent the current thinking of the Food and Drug Administration (FDA or Agency) on this topic. It does not establish any rights for any person and is not binding on FDA or the public. You can use an alternative approach if it satisfies the requirements of the applicable statutes and regulations. To discuss an alternative approach, contact the FDA staff or Office responsible for this guidance as listed on the title page.*

### I. Introduction

FDA developed this draft guidance document to provide clarity to industry and FDA staff about the current review practices for the Humanitarian Device Exemption (HDE) Program. This programmatic draft guidance addresses commonly asked questions about HDEs and Humanitarian Use Devices (HUDs), including FDA actions on HDE applications, post-approval requirements, and special considerations for devices marketed under the HDE Program. This draft guidance document reflects changes in the HDE Program resulting from statutory amendments made by the 21<sup>st</sup> Century Cures Act (Cures Act)<sup>1</sup> and explains the criteria FDA considers to determine if “probable benefit” has been demonstrated as part of the Agency’s decision-making process regarding marketing authorization for a HUD. This draft guidance document also reflects amendments made to the HDE provision of the Federal Food, Drug, and Cosmetic Act (FD&C Act) by the FDA Reauthorization Act of 2017 (FDARA).<sup>2</sup> Once finalized, this guidance will supersede “Guidance for HDE holders, Institutional Review Boards (IRBs), Clinical Investigators, and Food and Drug Administration Staff, Humanitarian Device Exemptions (HDE) Regulation: Questions and Answers,” issued July 8, 2010.

For the purposes of this guidance, “you” refers to the HDE applicant or holder, and “we” refers to FDA. FDA’s guidance documents, including this draft guidance, do not establish legally enforceable responsibilities. Instead, guidances describe the Agency’s current thinking on a topic and should be viewed only as recommendations, unless specific regulatory or statutory

<sup>1</sup> Pub. L. 114-255.

<sup>2</sup> Pub. L. 115-52.

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

101 requirements are cited. The use of the word *should* in Agency guidance means that something is  
102 suggested or recommended, but not required.

103 **II. Background**

104 HUDs are medical devices intended to benefit patients in the treatment or diagnosis of diseases  
105 or conditions that affect or are manifested in not more than 8,000 individuals in the United States  
106 per year.<sup>3</sup> In seeking marketing authorization under an HDE application, the first step is the  
107 preparation and submission of a HUD designation request to FDA’s Office of Orphan Products  
108 Development (OOPD). For more information on this step of the process, see the FDA guidance  
109 “[Humanitarian Use Device \(HUD\) Designations](#).”<sup>4</sup> We will refer to that document as the “HUD  
110 Designations Guidance” for the purposes of this draft guidance. The HDE application, which is  
111 the primary focus of this guidance document, is the second step in seeking marketing  
112 authorization for a HUD.

113 To the extent consistent with the protection of the public health and safety and with ethical  
114 standards, the purpose of the HDE provision is to “encourage the discovery and use of devices  
115 intended to benefit patients in the treatment and diagnosis of diseases or conditions that affect  
116 not more than 8,000 individuals in the United States,”<sup>5</sup>

117 FDA may grant an HDE, which is an exemption from the effectiveness requirements of sections  
118 514 and 515 of the FD&C Act, if we find that the device meets all of the following criteria:

- 119 1. The device will not expose patients to an unreasonable or significant risk of illness or  
120 injury, and the probable benefit to health from use of the device outweighs the risk of  
121 injury or illness from its use while taking into account the probable risks and benefits of  
122 currently available devices or alternative forms of treatment;
- 123 2. The device would not be available to a person with the disease or condition in question  
124 without the HDE, and no comparable device, other than another device approved under  
125 an HDE or Investigational Device Exemption (IDE)<sup>6</sup> is available to treat or diagnose  
126 such disease or condition; and
- 127 3. The device is designed to treat or diagnose a disease or condition that affects not more  
128 than 8,000 individuals in the United States on an annual basis.<sup>7</sup>

---

<sup>3</sup> 21 CFR 814.3(n). As subsequently explained, the current threshold is “not more than 8,000 individuals in the United States.” Before section 3052 of the Cures Act took effect, the threshold was “fewer than 4,000 individuals in the United States.”

<sup>4</sup> Available at <https://www.fda.gov/downloads/regulatoryinformation/guidances/ucm336515.pdf>.

<sup>5</sup> Section 520(m)(1) of the FD&C Act.

<sup>6</sup> An approved IDE permits a device to be shipped lawfully for the purposes of conducting investigations of the device without complying with certain other requirements of the FD&C Act that would apply to devices in commercial distribution. See section 520(g) of the FD&C Act; 21 CFR 812.1(a).

<sup>7</sup> See section 520(m)(2) of the FD&C Act; 21 CFR 814.104(b)(1)-(3).

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

129 The HDE provision was added to the FD&C Act by the Safe Medical Devices Act of 1990 and  
130 included, among other things, a prohibition on profits from sale of HUDs and a requirement that  
131 before “use” of a HUD to treat or diagnose patients at a facility, an IRB must approve such use.<sup>8</sup>  
132 For purposes of this draft guidance, approving the “use” of a HUD (as opposed to approving the  
133 “investigational use” or a “clinical investigation” of a device) refers to use of the HUD in the  
134 course of routine clinical care to treat or diagnose patients. Subsequent amendments to the  
135 FD&C Act have added important flexibility to the HDE program while retaining the purpose of  
136 encouraging the discovery of medical devices for use in limited patient populations.

137 The Food and Drug Administration Modernization Act of 1997 (FDAMA) included a section on  
138 expanding the humanitarian use of devices,<sup>9</sup> which among other provisions:

- 139 • Allowed for the use of HUDs under HDEs without prior IRB approval in certain  
140 emergency situations (see Section VIII.G for more information); and
- 141 • Provided that FDA may suspend or withdraw an HDE only after providing notice and an  
142 opportunity for an informal hearing (see Section VII.E for more information).

143 The Food and Drug Administration Amendments Act of 2007 (FDAAA) further modified the  
144 HDE provision, providing that HUDs indicated for use in pediatric patients, or in a pediatric  
145 subpopulation may be sold for profit, subject to certain restrictions.<sup>10</sup> The scope of HDE-  
146 approved devices eligible to make a profit was expanded by section 613 of the Food and Drug  
147 Administration Safety and Innovation Act (FDASIA).<sup>11</sup> Currently, a HUD is eligible to be sold  
148 for profit after receiving HDE approval only if, in addition to meeting other criteria, the number  
149 of HUDs sold during any calendar year does not exceed the annual distribution number (ADN),  
150 which is the number of devices reasonably needed to treat, diagnose, or cure a population of  
151 8,000 individuals in the United States (see Section VIII.A for more information).

152 FDAAA also added section 515A to the FD&C Act, which requires, among other things, the  
153 inclusion of additional information regarding pediatric uses in all original HDE applications, if  
154 such information is readily available.<sup>12</sup> Specifically, section 515A of the FD&C Act requires that  
155 each new HDE application include a description, based on readily available information, of any  
156 pediatric subpopulations that suffer from the disease or condition that the device is intended to  
157 treat, diagnose, or cure, and the number of affected pediatric patients.<sup>13</sup>

---

<sup>8</sup> Pub. L. 101-629, section 14.

<sup>9</sup> Pub. L. 105-115, section 203.

<sup>10</sup> Pub. L. 110-85, section 303. Pediatric patients are patients who are younger than 22 years of age at the time of diagnosis or treatment. See section 520(m)(6)(E)(i) of the FD&C Act; 21 CFR 814.3(s).

<sup>11</sup> Pub. L. 112-144. Many of the statutory provisions cited throughout this guidance, including sections 515A(a)(2) and 520(m)(6) of the FD&C Act, were added by section 302 of FDAAA and amended by FDASIA.

<sup>12</sup> See Pub L. 110-85, section 302. For further discussion of the information required for pediatric uses of medical devices under section 515A, see the guidance document, “Providing Information About Pediatric Uses of Medical Devices,” (available at <https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM339465.pdf>).

<sup>13</sup> See section 515A(a)(2) of the FD&C Act.

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

158 The Cures Act amended the FD&C Act to increase the maximum number of patients affected by  
159 the disease or condition that a HUD is designed to treat or diagnose to “not more than 8,000  
160 individuals in the United States.”<sup>14</sup> Further, the Cures Act removed the requirement that  
161 institutional review committees, i.e., IRBs, that supervise the clinical testing of devices or  
162 approve the use of HUDs be local.<sup>15</sup> See Sections VIII.E and VIII.F below for more information  
163 regarding differences between the “use” of a HUD and clinical investigations involving HUDs.  
164 The Cures Act also required FDA to publish a draft guidance that defines the criteria for  
165 establishing “probable benefit” as that term is used in section 520(m)(2)(C) of the FD&C Act.  
166 This draft guidance includes information to define those criteria for HDE applicants, other  
167 stakeholders, and FDA staff. See Section III for more information about the scope of this draft  
168 guidance.

169 FDARA further amended section 520(m)(4)(B) of the FD&C Act to allow either an IRB or “an  
170 appropriate local committee” to approve the use of a HUD to treat or diagnose patients at a  
171 facility.<sup>16</sup> We interpret this provision to provide additional flexibility for a healthcare facility to  
172 determine the individuals involved in, and processes and procedures used by, the committee that  
173 approves the use of HUDs at that facility in order to meet the needs of patients. Note, however,  
174 that this FDARA provision did not change the requirements for IRB oversight of a clinical  
175 investigation of a HUD. An “appropriate local committee” may not review and approve such a  
176 clinical investigation in place of an IRB.<sup>17</sup>

### 177 **III. Scope**

178 This draft guidance provides recommendations to industry and FDA staff about the operational  
179 aspects of the HDE Program and also explains the principal criteria that FDA considers when  
180 determining if probable benefit(s) to health have been demonstrated for a HUD that is being  
181 reviewed through the HDE Program. Additionally, this guidance addresses FDA’s assessment of  
182 whether the probable benefit(s) to health from use of the device outweighs the risk of injury or  
183 illness from its use, taking into account the probable risks and benefits of currently available  
184 devices or alternative forms of treatment. The decision tools in Appendices B and C are intended  
185 to help staff consider the probable benefit-risk factors discussed in Section VI when reviewing  
186 HDE applications. This guidance includes sections on the following topics, among others:

- 187 • FDA Review Actions for an HDE Application (Section V.B);
- 188 • Assessing Probable Benefit and Risk in an HDE Application<sup>18</sup> (Section VI);
- 189 • Post-Approval Requirements (Section VII);
- 190 • Special Considerations for Devices Marketed Under an HDE (Section VIII); and

---

<sup>14</sup> Pub. L. 114-255, section 3052.

<sup>15</sup> Pub. L. 114-255, section 3056.

<sup>16</sup> Pub. L. 115-52, section 502(b).

<sup>17</sup> See sections 520(g)(3) and 520(m)(4)(A) of the FD&C Act; 21 CFR part 56; 21 CFR part 812.

<sup>18</sup> As required by the Cures Act, this guidance explains the principal criteria that FDA considers when determining if probable benefit(s) to health have been demonstrated for a HUD that is being reviewed through the HDE Program.

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

- 191       • Appendices to support the HDE Program which include:  
192           ○ Filing Checklist (Appendix A)  
193           ○ Probable Benefit-Risk Assessment Tools (Appendices B and C)

194       The overarching principles in this draft guidance are applicable to devices that are eligible for  
195       review through an HDE application by CDRH as well as devices that are eligible for review  
196       through an HDE application by the Center for Biologics Evaluation and Research (CBER). This  
197       guidance is not intended to supplant or provide recommendations regarding device-specific data  
198       requirements, but it may cover broader areas not addressed in device-specific guidance  
199       documents. In addition, this guidance does not address review issues unique to combination  
200       products.

## 201       **IV. HUD Designations and HDE Applications**

202       Before submitting an HDE application to FDA, an HDE applicant must first prepare and submit  
203       a HUD designation request to OOPD and receive a HUD designation.<sup>19</sup> For more information on  
204       the preparation and submission of a HUD designation request, refer to 21 CFR 814.102(a) and  
205       the [HUD Designations Guidance](#). In the review of a HUD designation request, FDA will  
206       determine whether the device is for a rare disease or condition that affects or is manifested in not  
207       more than 8,000 individuals in the United States per year. In the case of a device used for  
208       diagnostic purposes, FDA will determine whether the documentation demonstrates that not more  
209       than 8,000 individuals per year would be subject to diagnosis by the device in the United  
210       States.<sup>20</sup> After receiving a HUD designation, the HDE applicant may submit an HDE application  
211       to the appropriate center. Each applicant must have its own HUD designation to submit an  
212       original HDE application for a proposed indication.<sup>21</sup> Additionally, the HDE applicant can utilize  
213       the HDE pathway only if no other comparable device (other than another device approved under  
214       an HDE or under an IDE) is available to treat or diagnose the disease or condition.

215       Note that if your device is part of a combination product, an HDE may not be the appropriate  
216       pathway to market. For questions about marketing pathways and regulatory requirements for  
217       combination products, please contact the Office of Combination Products by email at  
218       [combination@fda.gov](mailto:combination@fda.gov). For questions about Companion Diagnostic Devices, contact CDRH's  
219       Office of In Vitro Diagnostics and Radiological Health at [oir-policy@fda.hhs.gov](mailto:oir-policy@fda.hhs.gov).

## 220       **V. FDA's Review of HDE Applications**

221       Approval of an HDE application under 21 CFR part 814, Subpart H, is considered "FDA  
222       approval" of the device based on, among other criteria, evidence that the device will not expose  
223       patients to an unreasonable or significant risk of illness or injury and the probable benefit to  
224       health from use of the device outweighs the risk of injury or illness from its use, taking into

---

<sup>19</sup> 21 CFR 814.102(a).

<sup>20</sup> See 21 CFR 814.102(a)(5).

<sup>21</sup> See 21 CFR 814.102(a); 21 CFR 814.104(b)(1)

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

225 account the probable risks and benefits of currently available devices or alternative forms of  
226 treatment.<sup>22</sup> In addition, to be eligible for HDE application approval, FDA must determine that  
227 the device would not be available to a person with the disease or condition in question without  
228 the HDE application approval and that there is no comparable device, other than another device  
229 under an HDE or IDE, available to treat or diagnose the disease or condition.<sup>23</sup> A HUD that  
230 meets the HDE standard for approval is exempt from the requirement of establishing a  
231 reasonable assurance of effectiveness that would otherwise be required under sections 514 and  
232 515 of the FD&C Act, but is not exempt from the requirement for a reasonable assurance of  
233 safety.<sup>24</sup> FDA approval of an HDE application authorizes an applicant to market a HUD in  
234 accordance with approved labeling and indication(s) for use, subject to certain profit and use  
235 restrictions set forth in section 520(m) of the FD&C Act.

236 A HUD under an HDE may not serve as a predicate device for purposes of section 513(i) of the  
237 FD&C Act.<sup>25</sup> A Premarket Approval application (PMA) could subsequently be submitted for the  
238 same device and indication(s) approved under an HDE application if the applicant believes there  
239 is a reasonable assurance of safety and effectiveness; or, if appropriate, the applicant could  
240 instead submit a request for classification under section 513(f)(2) of the FD&C Act (a De Novo  
241 request). If FDA approves a PMA or grants a De Novo request for the HUD or another  
242 comparable device with the same indication, we may withdraw the HDE because the HUD  
243 would no longer meet the requirements of section 520(m)(2)(B) of the FD&C Act.<sup>26</sup> Section V.A  
244 discusses comparable devices, and Section VII.E discusses HUD designation re-evaluation.

245 FDA's review of HDE applications has similarities to the review of PMA applications, with a  
246 few key differences. Some similarities to the PMA program include:

- 247 • HDE amendments, supplements, and reports are generally subject to similar requirements  
248 as those for PMAs (although timeframes differ).<sup>27</sup> The requirements for each of these  
249 types of HDE submissions refers back to the regulatory requirements for the PMA  
250 counterpart.
- 251 • HUDs are subject to the quality system (QS) regulation under 21 CFR part 820, and HDE  
252 applications must include information in sufficient detail so that FDA can make a  
253 knowledgeable judgment about the quality control used in the manufacture of the

---

<sup>22</sup> Section 520(m)(2)(C) of the FD&C Act; 21 CFR 814.104(b)(3).

<sup>23</sup> Section 520(m)(2)(B) of the FD&C Act; 21 CFR 814.104(b)(2).

<sup>24</sup> See sections 514, 515, and 520(m) of the FD&C Act; 21 CFR 814.118(a)(1). As discussed in more detail in Section VI, FDA will perform an assessment of the probable benefits and risks of a device, considering a number of factors, including the target patient population and the size of the population, as well as available alternative treatments or diagnostics, as part of its determination of whether an HDE application meets the statutory standard for approval.

<sup>25</sup> Under 21 CFR 807.92(a)(3), a legally marketed (predicate) device to which a new device may be compared for a determination regarding substantial equivalence is a device that was legally marketed prior to May 28, 1976 (preamendments device), or a device which has been reclassified from class III to class II or I, or a device which has been found to be substantially equivalent through the 510(k) process.

<sup>26</sup> 21 CFR 814.118(a).

<sup>27</sup> See 21 CFR 814.106 (amendments); 814.108 (supplements); and 814.126 (reports).

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

254 device.<sup>28</sup> Additional information on manufacturing information to include in an HDE  
255 application can be found in the FDA guidance, “[Quality System Information for Certain](#)  
256 [Premarket Application Reviews](#).”<sup>29</sup> If you believe that you cannot comply with or should  
257 not be subject to the QS regulation requirements, you may request an exemption or a  
258 variance from any device QS regulation requirement. Petitions for an exemption or  
259 variance must be submitted according to the procedures set forth in 21 CFR 10.30.<sup>30</sup>

260 Key differences between the HDE and PMA programs include the following:

- 261 • A HUD under an HDE is *exempt* from the requirement of establishing a reasonable  
262 assurance of effectiveness that would otherwise be required under sections 514 and 515  
263 of the FD&C Act.
- 264 • HDE applications accepted for filing and to which the applicant does not submit a major  
265 amendment are reviewed in 75 days, rather than the traditional 180-day review timeframe  
266 for PMA applications.<sup>31</sup>
- 267 • HDE applications are not subject to user fees.
- 268 • For a device approved under an HDE application, medical device reports (MDRs)  
269 submitted to FDA in compliance with the requirements of 21 CFR part 803 shall also be  
270 submitted to the IRB of record.<sup>32</sup> If an appropriate local committee, instead of an IRB,  
271 approved the use of the device at a facility, FDA recommends that these MDRs be  
272 submitted to that committee. See Section VII below for additional information regarding  
273 IRB requirements.
- 274 • Use of HUDs in the clinical care of patients at a facility requires approval by either an  
275 IRB or an appropriate local committee, with the exception of emergency use.<sup>33</sup> See  
276 Section VII below for additional information regarding IRB and appropriate local  
277 committee requirements.
- 278 • An HDE holder can only make a profit, subject to the limit of the ADN, if the device  
279 meets the eligibility criteria for the exemption to the profit prohibition in section  
280 520(m)(6)(A)(i), subject to restrictions in section 520(m)(6) of the FD&C Act. See  
281 Section VIII.A below for additional information regarding eligibility for profit.
- 282 • The applicant must include in the application a statement that no comparable device,  
283 other than another HUD approved under an HDE or a device under an approved IDE, is  
284 available to treat or diagnose the disease or condition. The applicant must explain in its  
285 HDE application why the device would not be available for the indication in question  
286 without the HDE approval.<sup>34</sup>

---

<sup>28</sup> 21 CFR 814.104(b)(4).

<sup>29</sup> Available at <https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm070899.pdf>.

<sup>30</sup> See 21 CFR 820.1(e).

<sup>31</sup> 21 CFR 814.40 and 814.114.

<sup>32</sup> See 21 CFR 814.126(a).

<sup>33</sup> See section 520(m)(4) of the FD&C Act; 21 CFR 814.124(a).

<sup>34</sup> See section 520(m)(2)(B) of the FD&C Act and 21 CFR 814.104(b)(2).

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

287 Applicants wishing to submit a “modular HDE” for their HDE application may use the  
288 procedures outlined in the FDA guidance, “[Premarket Approval Application Modular Review](#).”<sup>35</sup>  
289 A modular HDE, for purposes of this draft guidance, is a compilation of sections or “modules”  
290 submitted at different times that together become a complete HDE application. HDE applicants  
291 should include a copy of or reference to FDA’s HUD designation letter with each HDE modular  
292 submission. The modules should conform to the information required for an HDE application.

293 When submitting an HDE application, applicants must prepare an electronic copy of their  
294 submission per the FDA guidance document, “[eCopy Program for Medical Device](#)  
295 [Submissions](#),”<sup>36</sup> and send the e-copy and cover letter to:

**For Products Regulated by CDRH**

U.S. Food and Drug Administration  
Center for Devices and Radiological Health  
Document Control Center – WO66-G609  
10903 New Hampshire Avenue  
Silver Spring, MD 20993-0002

**For Products Regulated by CBER**

U.S. Food and Drug Administration  
Center for Biologics Evaluation and Research  
Document Control Center – WO71-G112  
10903 New Hampshire Avenue  
Silver Spring, MD 20993-0002

296 **A. HDE Application Required Elements and Filing**  
297 **Review Principles**

298 To use this guidance appropriately, HDE applicants and FDA staff should review the following  
299 basic principles that are in bold typeface and followed by a description of FDA’s review policies  
300 and procedures. These principles, and the objective criteria outlined in the Filing Checklist in  
301 Appendix A, inform FDA’s HDE application filing decisions.

302 **The contents of the HDE application should allow the substantive review to proceed.**

303 The HDE regulations identify the criteria that, if not met, may serve as the basis for FDA  
304 refusing to file an HDE application.<sup>37</sup> The HDE application must contain the basic administrative  
305 and scientific elements listed in 21 CFR 814.104(b), unless the applicant justifies an omission in  
306 accordance with 21 CFR 814.104(c). The specific questions in the filing checklist are intended to  
307 help FDA ensure that the HDE application contents are sufficiently organized and complete to  
308 allow the review team to proceed with a substantive review of the application.

---

<sup>35</sup> Available at <https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM089767.pdf>.

<sup>36</sup> Available at <https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM313794.pdf>.

<sup>37</sup> 21 CFR 814.112(a).

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

309 **The filing decision should not be based on a substantive review of the data and information**  
310 **in the HDE application.**

311 The filing review is conducted to ensure that the HDE application is sufficiently complete and to  
312 determine the basic adequacy of the technical elements of the HDE application, respectively.  
313 Notably, in determining whether an HDE application should be filed, the submitted information  
314 should not be evaluated to determine whether the device will expose patients to an unreasonable  
315 risk of illness or injury or whether the probable benefit to health from the use of the device  
316 outweighs the risk of injury or illness from its use, taking into account the probable risks and  
317 benefits of currently available devices or alternative forms of treatment.

318 The checklist included in Appendix A is a tool to help ensure that the application contains the  
319 necessary information to conduct a substantive review. The elements in the checklist stem from  
320 either statutory or regulatory requirements, and the format and content are consistent with the  
321 analogous checklists for other types of premarket submissions. FDA generally should not refuse  
322 to file an HDE application because we have reviewed the data and believe that the application is  
323 ultimately inadequate to meet the standard for HDE approval. Subsequently, the substantive  
324 review of the HDE application will evaluate the quality of the content and lead to a decision  
325 regarding the safety and probable benefit of the HUD. Concerns identified by the Agency during  
326 the filing review regarding results and outcomes of nonclinical and clinical studies would not  
327 preclude filing.

328 **FDA should determine whether the applicant provided a justification for any alternative**  
329 **approach.**

330 If the applicant believes any criteria in the checklist are not applicable, the applicant should  
331 explain its rationale. Likewise, the applicant should provide a rationale for any deviation from an  
332 applicable device-specific or cross-cutting guidance document or FDA-recognized consensus  
333 standard. FDA expects that any item in the checklist that is missing from the application will be  
334 addressed with a rationale explaining why the item is not applicable and that any deviations will  
335 be explained. If a justification to omit certain information or for taking an alternative approach is  
336 provided, FDA will consider the adequacy of that justification or alternative approach during  
337 substantive review of the application. A given criterion in the checklist will be considered “Not  
338 Present” if the application fails to include either the information requested or a rationale for  
339 omission.

340 **HDE application filing reviews.**

341 FDA’s decision to “File” or “Not File” an HDE application should be made in collaboration with  
342 the HDE review team and with the appropriate supervisory oversight. FDA will notify the HDE  
343 applicant of the filing status within 30 calendar days from the date the HDE application was  
344 received.<sup>38</sup> Generally, a small number of missing items from the filing checklist will not preclude

---

<sup>38</sup> 21 CFR 814.112(a).

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

345 a positive filing decision; however, if multiple items are missing such that a substantive review  
346 cannot be completed, a “Not File” decision will typically be made.

#### 347 **Additional considerations when using the filing checklist.**

348 Certain elements of the HDE filing checklist are unique to the HDE Program. These elements are  
349 discussed in additional detail below.

#### 350 Amount charged for the device

351 As required by 21 CFR 814.104(b)(5), the applicant must state the amount to be charged for the  
352 device, and if the amount is more than \$250, a report or attestation must be provided verifying  
353 that the amount charged does not exceed the costs of the device’s research, development,  
354 fabrication, and distribution. A report must be prepared by an independent certified public  
355 accountant, made in accordance with the Statement on Standards for Attestation established by  
356 the American Institute of Certified Public Accountants. In lieu of such a report, an applicant may  
357 submit an attestation by a responsible individual of the organization, verifying that the amount  
358 charged does not exceed the costs of the device’s research, development, fabrication, and  
359 distribution. If the amount charged is \$250 or less, the requirement for a report by an  
360 independent certified public accountant or an attestation by a responsible individual of the  
361 organization is waived. Even if an HDE applicant requests that FDA consider whether the HUD  
362 meets certain eligibility criteria to qualify for profit making (see Section VIII.A. regarding how  
363 to request to make a profit), the applicant must still include this report or attestation in the HDE  
364 application.

#### 365 Comparable Devices

366 As required by 21 CFR 814.104(b)(2), the applicant must provide a statement that no other  
367 comparable device, other than another HUD approved under an HDE or a device under an  
368 approved IDE, is available to treat or diagnose the disease or condition. A “comparable device”  
369 does not need to be identical to the device that is the subject of the HDE application. However, in  
370 applying the “comparable device” exemption criterion, FDA takes into account that the purpose  
371 of the HDE Program is to encourage development of devices intended to treat or diagnose  
372 diseases or conditions that affect small patient populations. In determining whether a comparable  
373 device exists, FDA may consider:

- 374 • the device’s indications for use and technological characteristics;
- 375 • the patient population to be treated or diagnosed with the device; and
- 376 • whether the device meets the needs of the identified patient population.

377 FDA may refuse to file an HDE application if FDA determines that a comparable device is  
378 available (other than under another HDE or a device under an IDE) to treat or diagnose such

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

379 disease or condition for which the approval of the HDE application is being sought.<sup>39</sup> FDA  
380 cannot approve an HDE application for a HUD if we determine that such a comparable device is  
381 available.<sup>40</sup>

## 382 **B. FDA Review Actions for an HDE Application**

383 After an original HDE application or HDE supplement is accepted for filing and FDA begins its  
384 substantive review, the Agency may take the following actions during the course of review:<sup>41</sup>

- 385 • Approval order;
- 386 • Approvable letter;
- 387 • Major deficiency letter;
- 388 • Not approvable letter; and
- 389 • Denial order.

390 The review timeframe for original HDE applications and HDE supplements is 75 days.<sup>42</sup> In  
391 addition, if the applicant submits a major amendment, the review timeframe may be extended up  
392 to 75 days.<sup>43</sup> Certain changes to the manufacturing procedure or changes in method of  
393 manufacture may qualify to be submitted as a 30-day notice.<sup>44</sup> For more information regarding  
394 30-day notices, refer to the FDA guidance, “[30-Day Notices, 135-Day Premarket Approval  
395 \(PMA\) Supplements and 75-Day Humanitarian Device Exemption \(HDE\) Supplements for  
396 Manufacturing Method or Process Changes.](#)”<sup>45</sup>

### 397 **1. Approval Order**

398 FDA will issue an approval order (letter) informing the applicant that the HDE application is  
399 approved and that the applicant may begin commercial distribution of the device in accordance  
400 with any prescribed conditions of approval after we have completed our review and determined  
401 that none of the reasons listed in 21 CFR 814.118 for denying approval applies.<sup>46</sup>

402 When FDA issues an approval order, the FDA review clock stops. An approval order marks the  
403 end of FDA’s review, as it is a final action.

---

<sup>39</sup> See 21 CFR 814.112(a)(2).

<sup>40</sup> See section 520(m)(2)(B) of the FD&C Act.

<sup>41</sup> See 21 CFR 814.106 and 814.116.

<sup>42</sup> See 21 CFR 814.108 and 814.114.

<sup>43</sup> 21 CFR 814.106.

<sup>44</sup> See 21 CFR 814.108 and 814.39(f).

<sup>45</sup> Available at <https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM080194.pdf>.

<sup>46</sup> See 21 CFR 814.116(b).

## 2. Approvable Letter

404  
405 FDA will issue an approvable letter to inform the applicant that we have completed our review of  
406 the application and determined that one or both of the following are necessary:<sup>47</sup>

- 407 • Resolution of minor deficiencies that are identified in the approvable letter. These  
408 deficiencies may include, for example, clarifications of previously submitted information,  
409 revisions to the labeling, and revisions to or development of a post-approval study  
410 protocol.
- 411 • Completion of an FDA inspection that finds the manufacturing facilities, methods, and  
412 controls in compliance with the QS regulation, 21 CFR part 820, and, if applicable,  
413 verifies records pertinent to the HDE application. When this is the case, the approvable  
414 letter states that the device is “approvable pending GMP inspection.”

415 When FDA issues an approvable letter pending resolution of minor deficiencies, we stop the  
416 FDA review clock and place the application on hold. When FDA receives a complete response to  
417 an approvable letter for an HDE application, we consider it a major amendment and restart the  
418 clock with a new 75-day FDA response timeframe.

419 When FDA issues an approvable letter pending GMP inspection letter, we stop the FDA review  
420 clock. Once FDA determines that the GMP issues are resolved, we will issue an approval order if  
421 all other minor deficiencies that may have been noted in the approvable letter have also been  
422 resolved.

## 3. Major Deficiency Letter

423  
424 FDA will issue a major deficiency letter to inform the applicant that the HDE application lacks  
425 significant information necessary for FDA to complete our review and request that the applicant  
426 amend the application to provide the necessary information regarding the device,<sup>48</sup> such as:

- 427 • Additional clinical experience to demonstrate safety and probable benefit;
- 428 • Additional non-clinical data to demonstrate safety and probable benefit of the device  
429 (e.g., electromagnetic compatibility, electrical safety, biocompatibility, reliability,  
430 software, labeling, animal testing<sup>49</sup>);
- 431 • Scientific rationale for test data provided in the application;
- 432 • New validation data and analyses (e.g., due to device modifications made during the  
433 course of the HDE application review); or

---

<sup>47</sup> See 21 CFR 814.116(c).

<sup>48</sup> See 21 CFR 814.106 and 21 CFR 814.37(b).

<sup>49</sup> We support the principles of the “3Rs,” to reduce, refine, and replace animal use in testing when feasible. We encourage sponsors to consult with us if they wish to use a non-animal testing method they believe is suitable, adequate, validated, and feasible. We will consider if such an alternative method could be assessed for equivalency to an animal test method.

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

- 434       • A re-analysis of previously submitted data (e.g., alternative analytical method).

435       When FDA issues a major deficiency letter, we stop the FDA review clock and place the  
436       application on hold. When FDA receives a complete response to a major deficiency letter, we  
437       consider it a major amendment, restart the clock, and review the amendment within 75 days.

## 438                   **4. Not Approvable Letter**

439       FDA will issue a not approvable letter to inform the applicant that we have completed our review  
440       and that we do not believe that the application can be approved ‘as-is’ because of significant  
441       deficiencies. The not approvable letter will describe the deficiencies in the application, including  
442       each applicable ground for not approving and, where practical, will identify measures required to  
443       place the application in an approvable form.<sup>50</sup>

444       Generally, before FDA issues a not approvable letter, we will first issue a major deficiency letter  
445       to provide the applicant with an opportunity to address concerns. However, if an applicant fails  
446       to provide an adequate response to a major deficiency letter, or if we have attempted to resolve  
447       all deficiencies via interactive review and have not received adequate responses, FDA will issue  
448       a not approvable letter.

449       When FDA issues a not approvable letter, we stop the review clock and place the application on  
450       hold. If FDA receives a complete response to a not approvable letter, the amendment will be  
451       considered a major amendment, and we restart the clock with a new 75-day FDA response  
452       timeframe.<sup>51</sup>

## 453                   **5. Denial Order**

454       FDA may deny approval of an HDE application for any of the reasons identified in 21 CFR  
455       814.118(a). FDA will issue a denial order (letter) when we need to inform the applicant that we  
456       have denied approval of the HDE application. The denial order will identify all deficiencies in  
457       the application, including each applicable ground for denial and, where practical, will identify  
458       measures required to place the application in an approvable form. The denial order will include a  
459       notice of an opportunity to request review under section 515(d)(4) of the FD&C Act.<sup>52</sup>

460       When FDA issues a denial order, we end the FDA review clock if a prior action has not already  
461       done so. FDA expects that a denial will normally be preceded by another FDA action that stops  
462       the review clock, such as a not approvable letter. However, the FD&C Act does not require any  
463       prior FDA action, and FDA may, in appropriate circumstances, proceed directly to issuing a  
464       denial order. A denial order marks the end of FDA’s review, as it is considered a final action.

---

<sup>50</sup> 21 CFR 814.116(d).

<sup>51</sup> See 21 CFR 814.44(f)(1) and 814.116(d).

<sup>52</sup> 21 CFR 814.118(b) and 814.45(b).

465

## **6. Acknowledgement of Voluntary Withdrawal**

466 Under FDA regulations for review of HDE applications, FDA considers an original HDE  
467 application or HDE supplement to have been voluntarily withdrawn if an applicant fails to  
468 respond to an approvable letter, major deficiency letter, or not approvable letter within 75 days  
469 of issuance of the letter.<sup>53</sup> However, if before the end of that 75-day period, an HDE applicant  
470 requests additional time to generate data or provide other information to address the issues  
471 identified in the FDA letter, FDA may agree to allow additional time, as appropriate. When  
472 additional time is requested, FDA generally would allow up to 360 days to provide a complete  
473 response to the FDA action letter. We generally do not find it appropriate to agree to requests for  
474 additional time beyond 360 days. FDA intends to notify the applicant with a letter  
475 acknowledging voluntary withdrawal of the HDE application or HDE supplement, and any  
476 amendment submitted in response to an FDA action letter after FDA’s notification  
477 acknowledging voluntary withdrawal would be considered a resubmission of the HDE  
478 application. As such, it will be assigned a new HDE number and will be subject to the  
479 requirements of 21 CFR 814.104.

## **VI. Assessing Probable Benefit and Risk in an HDE Application**

482 As discussed above, a device that meets the criteria under section 520(m) of the FD&C Act is  
483 exempt from the effectiveness requirements of sections 514 and 515 of the FD&C Act but is not  
484 exempt from the requirement for a reasonable assurance of safety.<sup>54</sup> To approve an HDE  
485 application, section 520(m) of the FD&C Act requires that FDA find, among other things, “that  
486 the device will not expose patients to an unreasonable or significant risk of illness or injury and  
487 the probable benefit to health from the use of the device outweighs the risk of injury or illness  
488 from its use, taking into account the probable risks and benefits of currently available device[s]  
489 or alternative forms of treatment.”<sup>55</sup> To make these findings and evaluate safety and probable  
490 benefit, FDA performs an assessment of probable risks and benefits for a device as part of its  
491 HDE application review.

492 FDA also assesses probable benefits and risks as part of its review of PMAs and De Novo  
493 requests,<sup>56</sup> and the Agency has previously presented a benefit-risk framework for benefit-risk

---

<sup>53</sup> See 21 CFR 814.106.

<sup>54</sup> See sections 514, 515, and 520(m) of the FD&C Act; 21 CFR 814.118(a)(1). FDA regulations state that there is a reasonable assurance that a device is safe “when it can be determined, based upon valid scientific evidence, that the probable benefits to health from use of the device for its intended uses and conditions of use, when accompanied by adequate directions and warnings against unsafe use, outweigh any probable risks.” 21 CFR 860.7(d)(1).

<sup>55</sup> Section 520(m)(2)(C) of the FD&C Act.

<sup>56</sup> See FDA’s guidance, “Factors to Consider When Making Benefit-Risk Determinations in Medical Device Premarket Approval and De Novo Classifications,” (<https://www.fda.gov/downloads/medicaldevices/deviceregulationandguidance/guidancedocuments/ucm517504.pdf>) for a discussion of how FDA benefit-risk

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

494 determinations in the context of reviewing those applications in the guidance document “[Factors](#)  
495 [to Consider When Making Benefit-Risk Determinations in Medical Device Premarket Approval](#)  
496 [and De Novo Classifications](#).” FDA believes that the benefit-risk framework and factors used to  
497 assess PMAs or De Novo requests referenced above is generally appropriate for HDE  
498 applications. FDA therefore intends to consider the same factors described in FDA’s benefit-risk  
499 framework for evaluating PMAs or De Novo requests when assessing probable benefits and risks  
500 for HDE applications.

501 However, given the different standards and requirements that apply to approval of an HDE  
502 application, the weighting of those factors and the nature of the evidence available regarding  
503 those factors is likely to differ in the HDE context. Among other differences, a reasonable  
504 assurance of effectiveness is not required for a device approved under an HDE application.<sup>57</sup>  
505 Therefore, when compared to a PMA or De Novo request, both of which require a demonstration  
506 of reasonable assurance of safety and effectiveness,<sup>58</sup> there is generally likely to be greater  
507 uncertainty surrounding the benefit-risk profile based on the evidence submitted in an HDE  
508 application.

509 Moreover, as with the benefit-risk framework for evaluating PMAs or De Novo requests, FDA  
510 considers relevant factors as part of the probable benefit-risk assessment for an HDE application  
511 in the context of the intended use of the device, including the target patient population and the  
512 size of the population. FDA’s probable benefit-risk assessment also takes into account currently  
513 available alternative treatments or diagnostics, including their limitations. When available,  
514 information characterizing patients’ tolerance for risk and their perspective on probable benefit  
515 may provide useful context during this assessment. Refer to the FDA guidance document,  
516 “[Patient Preference Information—Voluntary Submission, Review in Premarket Approval](#)  
517 [Applications, Humanitarian Device Exemption Applications, and De Novo Requests, and](#)  
518 [Inclusion in Decision Summaries and Device Labeling](#).”<sup>59</sup>

519 The Agency’s probable benefit-risk framework provides for flexibility and use of scientific  
520 judgment in assessing the totality of the evidence to determine if a specific device meets the  
521 standard for HDE application approval. This flexibility allows FDA to take into account  
522 considerations relevant to HDE applications (e.g., a relatively small patient population) under a  
523 framework that is consistent across device marketing submission types. To do so, FDA has  
524 developed tools to assist in assessing probable benefit and risk for an HDE application. Refer to  
525 the supplementary Considerations for the Probable Benefit-Risk Assessment in Appendix B and

---

assessments relate to the Agency’s safety and effectiveness evaluation of a device in the context of reviewing PMAs and De Novo requests.

<sup>57</sup> Section 520(m)(2) of the FD&C Act.

<sup>58</sup> To meet the statutory standard for approval of a PMA, there must be a showing of reasonable assurance that the device is safe and effective. See section 515(d) of the FD&C Act. The De Novo classification process is appropriate for devices that would otherwise be subject to PMA but for which general controls or general and special controls provide a reasonable assurance of safety and effectiveness. See section 513(f)(2) of the FD&C Act.

<sup>59</sup> Available at <https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM446680.pdf>.

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

526 the Probable Benefit-Risk Assessment Tool in Appendix C. These tools are intended to reflect  
527 differences in probable benefit-risk determinations for an HDE application when compared to  
528 other types of device marketing submissions. Note that the tools in this draft guidance also  
529 present questions to consider regarding the factors for the probable benefit-risk assessment  
530 differently than the benefit-risk guidance for PMAs and De Novos. While the factors have the  
531 same meaning, the tools for HDE applications have been reorganized to assist in weighing the  
532 probable benefits and risks.

533 FDA has also published guidance with respect to making benefit-risk determinations for IDE  
534 applications.<sup>60</sup> However, unlike an IDE, which permits a device to be shipped lawfully for the  
535 purpose of conducting a clinical investigation of the device’s safety and/or effectiveness, an  
536 approved HDE application is a marketing authorization. Accordingly, the two applications have  
537 different statutory and regulatory standards and, as a general matter, earlier stages of device  
538 development and investigational study under an IDE are typically associated with greater  
539 uncertainty than an HDE. Approval of an IDE application to permit investigational use of a  
540 device may be appropriate where it is unknown if subjects are likely to benefit from the use of  
541 the device, if the risks to the subjects are outweighed by the anticipated benefits to the subjects  
542 and the importance of the knowledge to be gained, and the IDE application otherwise satisfies  
543 the requirements of 21 CFR part 812.<sup>61</sup> In contrast, approval of the HDE application, which  
544 authorizes the marketing of a device, requires, among other things, a demonstration that there is  
545 probable benefit and that the probable benefit outweighs the risk of injury or illness from its use,  
546 taking into account the probable risks and benefits of currently available devices or alternative  
547 forms of treatment.

548 For HDE applications, probable benefit is present when there is evidence for FDA to reasonably  
549 conclude that patients are likely to benefit from the use of the device. The probable benefit-risk  
550 decision support tools prompt FDA review staff to consider probable benefit in terms of:

- 551 • Type of benefit(s)
- 552 • Magnitude of benefit(s);
- 553 • Probability of the patient experiencing one or more benefit(s);
- 554 • Duration of effect(s); and
- 555 • Patient perspectives.

556 In addition, the probable benefit-risk decision support tools prompt FDA review staff to consider  
557 risk in terms of:

- 558 • Severity, types, number and rates of harmful events associated with use of the device;

---

<sup>60</sup> See “Factors to Consider When Making Benefit-Risk Determinations for Medical Device Investigational Device Exemptions,” (<https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM451440.pdf>).

<sup>61</sup> FDA may disapprove or withdraw approval of an IDE application if, among other reasons, “[t]here is reason to believe that the risks to the subjects are not outweighed by the anticipated benefits to the subjects and the importance of the knowledge to be gained . . . .” 21 CFR 812.30(b)(4).

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

- 559 • Probability of a harmful event;
- 560 • Duration of harmful events;
- 561 • Risk from false-positive or false-negative results for diagnostics; and
- 562 • Patient perspective.

563 As with benefit-risk assessments for PMAs and De Novo requests, FDA considers additional  
564 factors, including uncertainty and available alternative treatments or diagnostics, as part of  
565 assessing whether probable benefits outweigh the probable risks in the context of an HDE  
566 application. Sources of evidentiary uncertainty could include, but are not limited to:

- 567 • Sample size;
- 568 • Duration of follow-up;
- 569 • Use of a surrogate outcome; and/or
- 570 • Use of non-clinical performance data such as animal testing or computer modeling rather  
571 than or in addition to a clinical or surrogate outcome.

572 FDA recognizes that in some instances there may be little or no clinical experience with the  
573 device that is the subject of an HDE application. Depending upon the nature of the device and its  
574 associated risks, FDA may request that clinical data be collected in support of an HDE  
575 application. However, it is also important to recognize that non-clinical data may play a critical  
576 role in probable benefit-risk assessments in the context of HDE. Medical devices often have  
577 attributes that cannot be tested by clinical methods alone and that play a major role in the  
578 performance, safety, or effectiveness of the device. In some cases, non-clinical testing (e.g.,  
579 engineering performance studies, animal studies, analytical testing, or computer modeling and  
580 simulation) can obviate or reduce the need for clinical testing to evaluate certain aspects of  
581 device design or performance. Both clinical and non-clinical testing methods may be used to  
582 assess the probable benefit (including consideration of its likelihood, magnitude and duration),  
583 the probability or severity of a given risk, and/or the success of risk control measures, including  
584 risk mitigation measures.

585 The document at Appendix B, “Considerations for the Probable Benefit-Risk Assessment,” is  
586 intended to complement the Probable Benefit-Risk Assessment Tool at Appendix C. When  
587 finalized, FDA staff should use the two tools together while reviewing the HDE application.  
588 FDA staff should also refer to this guidance, when finalized, to assist them in making their  
589 determinations. HDE applicants may also consider these tools, but inclusion of these documents  
590 as part of an application is not a requirement.

## 591 **VII. Post-Approval Requirements**

592 HDE applications are subject to a number of post-approval requirements, as described below. In  
593 addition to these requirements, a post-approval study (PAS) may be required and described in the

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

594 approval order.<sup>62</sup> The guidance document, “[Procedures for Handling Post-Approval Studies](#)  
595 [Imposed by PMA Order](#),” includes recommendations for PMA post-approval studies.<sup>63</sup>  
596 However, most of the information in this guidance document also applies to post-approval  
597 studies imposed for an approved HDE application.

## 598 **A. IRB or Appropriate Local Committee Oversight** 599 **and Approval**

600 A HUD with an approved HDE application is approved by FDA for marketing. However, the  
601 HDE holder is responsible for ensuring that a HUD under an HDE is administered only in  
602 facilities having IRB oversight in accordance with the Agency’s regulation governing IRBs.<sup>64</sup> In  
603 addition, approval by an IRB or an appropriate local committee is required before a HUD under  
604 an HDE can be used at a facility for clinical care, with the exception of emergency use.<sup>65</sup> See  
605 Section VIII.G., “Emergency Use of HUDs.” Note that an IRB or appropriate local committee is  
606 not required to review and approve each individual use of a HUD and may grant a generalized  
607 approval to use the HUD at a facility. In such circumstances, FDA does not require the facility,  
608 HDE holder, or practitioner to seek approval from the IRB or appropriate local committee for  
609 each use, provided the use of the HUD is within the terms of the generalized approval.

610 The HDE holder is not required to submit the names and addresses of the reviewing IRBs or  
611 appropriate local committees to FDA. However, as required in 21 CFR 814.126(b)(2), the HDE  
612 holder must maintain records of:

- 613 • The names and addresses of the facilities to which the HUD was shipped;
- 614 • Correspondence with reviewing IRBs; and
- 615 • Any other information requested by a reviewing IRB or FDA.

616 FDA recommends that HDE holders likewise maintain correspondence with reviewing  
617 appropriate local committees as well as other information that such committees may require.  
618 Additional information regarding the role of IRBs and appropriate local committees with respect  
619 to HUDs is available in Section VIII.E, below.

## 620 **B. Adverse Event Reporting**

621 Whether expected or not, adverse events must be reported and evaluated in accordance with  
622 Medical Device Reporting requirements in 21 CFR part 803. Device manufacturers and user  
623 facilities are required to submit medical device reports to FDA and to the “IRB of record” (i.e.,  
624 the IRB that oversees use of the HUD at the facility where the adverse event occurred) after

---

<sup>62</sup> 21 CFR 814.126(a) and 814.82(a)(2).

<sup>63</sup> Available at <https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm071013.pdf>.

<sup>64</sup> See section 520(m)(4)(A) of the FD&C Act. FDA regulations governing IRBs are in 21 CFR part 56.

<sup>65</sup> See section 520(m)(4)(B) of the FD&C Act and 21 CFR 814.124(a).

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

625 HDE approval.<sup>66</sup> In the event that an appropriate local committee approved the use of the HUD  
626 for routine clinical care at that facility, instead of an IRB, we recommend that manufacturers  
627 submit MDRs to that committee.

628 Accordingly, manufacturers must submit MDRs to FDA and the IRB of record when they  
629 become aware of information reasonably suggesting that a HUD may have caused or contributed  
630 to a death or serious injury, or has malfunctioned and would be likely to cause or contribute to a  
631 death or serious injury if the malfunction were to recur.<sup>67</sup>

632 User facilities must submit MDRs to FDA, the IRB of record, and the manufacturer when they  
633 become aware of information reasonably suggesting that a HUD may have caused or contributed  
634 to the death of a patient, and must submit reports to the manufacturer (or to FDA and the IRB of  
635 record if the manufacturer is unknown) when they become aware of information reasonably  
636 suggesting that a HUD may have caused or contributed to a serious injury to a patient of the  
637 facility.<sup>68</sup> As defined by 21 CFR 803.3(w), a serious injury means an injury or illness that:

- 638 • Is life-threatening;
- 639 • Results in permanent impairment of a body function or permanent damage to a body  
640 structure; or
- 641 • Necessitates medical or surgical intervention to preclude permanent impairment of a  
642 body function or permanent damage to a body structure.

643 Pediatric adverse events will be reviewed periodically by FDA's Pediatric Advisory Committee  
644 (PAC).<sup>69</sup>

645 If the HUD is being investigated in a clinical study under an IDE, adverse events that occur  
646 during the study should be reported in accordance with 21 CFR 812.150(a)(1) and 812.150(b)(1).

### **C. HDE Supplements**

648 After FDA approval of an original HDE application, an applicant shall submit an HDE  
649 supplement for review and approval by FDA before making a change affecting the safety or  
650 probable benefit of the device.<sup>70</sup>

651 If you wish to request new indications for use for a device under an HDE (e.g., for a different  
652 disease or condition) you must obtain a new HUD designation and submit a new original HDE  
653 application in compliance with 21 CFR 814.110. If you are submitting a new original HDE  
654 application, please contact OOPD to discuss obtaining a new HUD designation. In the new HDE

---

<sup>66</sup> See section 519(a) and (b) of the FD&C Act; 21 CFR 803.30, 803.50, and 814.126(a). See 21 CFR 803.3(d) for the definition of a device user facility.

<sup>67</sup> See 21 CFR 803.50.

<sup>68</sup> 21 CFR 803.30.

<sup>69</sup> Section 520(m)(8) of the FD&C Act.

<sup>70</sup> 21 CFR 814.108 and 814.39.

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

655 application, any relevant information or data submitted in the HDE application for the original  
656 indication may be incorporated by reference.

657 **D. HDE Periodic Reports**

658 You must submit periodic reports for HDEs in accordance with the approval order under 21 CFR  
659 814.126(b). HDE periodic reports must include the following information unless FDA specifies  
660 otherwise:

- 661 • An update of the information required under 21 CFR 814.102(a) to demonstrate that the  
662 HUD designation is still valid.<sup>71</sup> An updated annual incidence reassessment (AIR) based  
663 on updated numbers to show that the target population for the disease or condition for  
664 which the device has been designated is not more than 8,000 per year provides this  
665 information. The AIR refers to the HUD designated population, which in some cases may  
666 be larger than the approved indication under the HDE (i.e., if the HDE approval covers  
667 only a certain indication within the designated disease or condition). In reviewing this  
668 information, the reviewing center, CDRH or CBER, may refer the AIR to OOPD for  
669 further evaluation if necessary.
- 670 • An update to the explanation of why the device would not be available unless an HDE  
671 were granted, a statement that no other comparable device (other than another HUD  
672 under an HDE or a device under an IDE) is available to treat or diagnose the disease or  
673 condition, and an updated discussion of the risks and benefits of currently available  
674 devices or alternative forms of treatment in the United States.<sup>72</sup>
- 675 • An update to the explanation of why the probable benefits to health from the use of the  
676 device outweighs the risk of injury or illness from its use, taking into account the  
677 probable risks and benefits of currently available devices or alternative forms of  
678 treatment.<sup>73</sup>
- 679 • An update to the amount to be charged for the device and, if over \$250, a report by an  
680 independent certified public accountant or an attestation by a responsible individual of  
681 the organization verifying that the amount charged does not exceed the costs of the  
682 device's research, development, fabrication, and distribution.<sup>74</sup>
- 683 • The number of devices that have been shipped or sold since initial marketing approval  
684 and, if the number shipped or sold exceeds 8,000, an explanation and estimate of the  
685 number of devices used per patient.<sup>75</sup> FDA interprets this regulation to require HDE  
686 holders to report the total number of HUDs shipped or sold pursuant to the approved  
687 HDE application, no matter how the HUDs are used (whether for the approved  
688 indication(s), emergency use, or otherwise). However, for devices that have both an HDE  
689 application approval and a PMA approval for a different indication, you need report only

---

<sup>71</sup> See 21 CFR 814.126(b)(1)(i).

<sup>72</sup> See 21 CFR 814.126(b)(1)(ii) and 814.104(b)(2).

<sup>73</sup> See 21 CFR 814.126(b)(1)(ii) and 814.104(b)(3).

<sup>74</sup> See 21 CFR 814.126(b)(1)(ii) and 814.104(b)(5).

<sup>75</sup> See 21 CFR 814.126(b)(1)(iii).

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

- 690 the number of devices that are shipped or sold pursuant to the HDE unless specifically  
691 required otherwise by the PMA Approval Order.
- 692 • Information describing the clinical experience with the approved device, including safety  
693 information that is known or reasonably should be known to the applicant, and any  
694 medical device report made under 21 CFR part 803.<sup>76</sup>
  - 695 • A summary of any changes made to the device in accordance with supplements submitted  
696 under 21 CFR 814.108.<sup>77</sup>

697 **E. HUD Designation Re-Evaluation**

698 If, based on information contained in the HDE periodic reports, FDA is concerned that the HUD  
699 designation may no longer apply to your device, we may contact you for additional information,  
700 re-evaluate, and possibly revoke your HUD designation.<sup>78</sup>

701 If we make the determination that more than 8,000 individuals in the United States are affected  
702 by or manifest a certain disease or condition per year, we may consider whether your HUD  
703 designation should be revoked in accordance with 21 CFR 814.102(c) and your HDE withdrawn  
704 in accordance with 21 CFR 814.118. In making this determination, we intend to consider factors  
705 such as the number of patients with the disease or condition and the public health need for the  
706 device. We intend to discuss the regulatory options with the HDE holder before revoking a HUD  
707 designation. The investigational use of a HUD outside of the HUD designation would not count  
708 toward the limit of 8,000 individuals per year.

709 If FDA subsequently approves a PMA or grants a De Novo request for the HUD or another  
710 comparable device with the same indication(s), we may withdraw the HDE because the HUD  
711 would no longer meet the requirements of section 520(m)(2)(B) of the FD&C Act.<sup>79</sup>

712 **VIII. Special Considerations for Devices Marketed**  
713 **Under an HDE**

714 **A. Eligibility for Profit**

715 Except in certain circumstances, HUDs under an HDE cannot be sold for an amount that exceeds  
716 the costs of research and development, fabrication, and distribution of the device (i.e., for profit).  
717 If a HUD is studied in a clinical investigation for a new indication, the sponsor of the clinical  
718 investigation cannot charge subjects or investigators a price higher than necessary to recover the  
719 costs of manufacture, research, development, and handling.<sup>80</sup> Any costs for which a subject in a

---

<sup>76</sup> See 21 CFR 814.126(b)(1)(iv).

<sup>77</sup> See 21 CFR 814.126(b)(1)(v).

<sup>78</sup> See 21 CFR 814.102(c) and 814.126(b)(1).

<sup>79</sup> See 21 CFR 814.118(a).

<sup>80</sup> See 21 CFR 812.7(b).

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

720 clinical investigation is responsible must, when appropriate, be provided in the informed consent  
721 document.<sup>81</sup>

722 Under section 520(m)(6)(A)(i) of the FD&C Act, as amended by FDASIA, a HUD under an  
723 HDE is only eligible to be sold for profit if the device meets the following criteria (i.e., the  
724 eligibility criteria):

- 725 • The device is intended for the treatment or diagnosis of a disease or condition that occurs  
726 in pediatric patients or in a pediatric subpopulation, and such device is labeled for use in  
727 pediatric patients or in a pediatric subpopulation in which the disease or condition occurs;  
728 or
- 729 • The device is intended for the treatment or diagnosis of a disease or condition that does  
730 not occur in pediatric patients or that occurs in pediatric patients in such numbers that the  
731 development of the device for such patients is impossible, highly impracticable, or  
732 unsafe.

733 If a device under an HDE does not meet either of the eligibility criteria, the device cannot be sold  
734 for profit. FDA reviews the financial information in the HDE holder's initial application and  
735 periodically thereafter. When approving the use of a HUD for treatment or diagnosis of patients  
736 in clinical care, the IRB or appropriate local committee is not required to review a justification  
737 for the amount charged for the HUD. The descriptions below are intended to provide additional  
738 clarity regarding each component of the eligibility criteria.

739 **Occurs in pediatric patients or in a pediatric subpopulation** – This would be a disease  
740 or condition that occurs in patients who are younger than 22 years of age.

741 **Does not occur in pediatric patients** – This would be a disease or condition that occurs  
742 only in patients who are 22 years of age or older. An example of a disease that does not  
743 occur in pediatric patients is Alzheimer's disease.

744 **Impossible or highly impracticable** – When determining whether the development of a  
745 HUD in pediatric patients is "impossible" or "highly impracticable," FDA considers  
746 information provided by the applicant to FDA, including publicly-available information  
747 such as published literature, which demonstrates that the sponsor would not be able to  
748 conduct the necessary clinical investigation(s) in the pediatric population for the device.

749 For example, FDA may determine that the development of a particular device is  
750 "impossible" or "highly impracticable" in pediatric patients if the device is intended to  
751 treat a disease or condition that has a pediatric annual incidence that is so small, or if the  
752 prevalence of the pediatric patients living with the disease is so small, or if the pediatric  
753 population is so geographically dispersed to prevent sufficient patient recruitment in the  
754 pediatric population for a clinical investigation. Because of the speed and efficiency of

---

<sup>81</sup> 21 CFR 50.25(b)(3).

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

755 modern communications tools, geographic dispersion will justify a waiver only in  
756 extraordinary circumstances and will generally have to be coupled with very small  
757 population size. FDA does not consider economic factors (such as the costs associated  
758 with conducting a clinical investigation) as a basis for being “impossible” or “highly  
759 impracticable.”

760 **Unsafe** – FDA may determine that the development of a HUD in pediatric patients is  
761 “unsafe” if the applicant has provided information, including publicly-available  
762 information such as published literature, to FDA that demonstrates that the device would  
763 expose pediatric patients to an unreasonable or significant risk of illness or injury. If FDA  
764 determines that the HUD is eligible to be sold for profit because development of the  
765 device in pediatric patients would be “unsafe,” the labeling (e.g., warnings or  
766 contraindications) for the device should reflect the safety concern.

767 An HDE applicant whose device meets one of the eligibility criteria and who wishes to sell its  
768 HUD for profit should provide adequate supporting documentation to FDA in its original HDE  
769 application to demonstrate to FDA that the HUD meets the eligibility criteria. An HDE holder  
770 whose HDE application was approved prior to the enactment of FDASIA on July 9, 2012, and  
771 who wishes to sell its HUD for profit should provide adequate supporting documentation to FDA  
772 in an HDE supplement to demonstrate to FDA that the HUD meets the eligibility criteria. If FDA  
773 determines that the HUD meets the eligibility criteria, FDA will then determine the ADN for the  
774 HUD when FDA approves the HDE application or supplement.<sup>82</sup>

## 775 **B. The Annual Distribution Number (ADN)**

776 Under section 520(m)(6) of the FD&C Act, if FDA makes a determination that a HUD meets the  
777 eligibility criteria, you may sell the HUD for profit after receiving HDE application approval as  
778 long as the number of devices distributed in any calendar year does not exceed the ADN for the  
779 device.

780 The ADN is determined by FDA:

- 781 • When the Agency approves the original HDE application; or
- 782 • When the Agency approves an HDE supplement for an HDE application approved before  
783 the enactment of FDASIA on July 9, 2012, if the HDE holder seeks a “determination” for  
784 the HUD in an HDE supplement based upon the profit-making eligibility criteria, and  
785 FDA determines that the HUD meets the eligibility criteria.<sup>83</sup>

786 Under section 520(m)(6)(A)(ii) of the FD&C Act, the ADN is defined, with respect to a device  
787 under an HDE, as the number of devices “reasonably needed to treat, diagnose, or cure a  
788 population of 8,000 individuals in the United States.” The applicant should provide the number

---

<sup>82</sup> See section 613(b) of FDASIA and Section VIII.B. for more discussion on the ADN.

<sup>83</sup> See section 520(m)(6)(A)(ii) of the FD&C Act and section 613(b) of FDASIA.

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

789 of devices per year reasonably needed for each individual in the HDE application or HDE  
790 supplement and provide adequate supporting documentation to support such number in order to  
791 provide a basis for FDA to calculate the ADN.

792 When determining the ADN, FDA considers the number of devices per year reasonably needed  
793 to treat, diagnose, or cure an individual (“first multiplier”) and multiplies that value by 8,000  
794 (“second multiplier”). By law, the second multiplier is always 8,000. Therefore, the ADN will be  
795 equal to or greater than 8,000, depending on the value of the first multiplier. For example, the  
796 target population estimate for the intended use may be 3,000 individuals, but if 2 devices are  
797 reasonably needed per year to treat, diagnose or cure a patient, the ADN would be 16,000 (i.e., 2  
798 multiplied by 8,000 because the second multiplier for the ADN is always 8,000, regardless of the  
799 actual population estimate). After FDA has determined the ADN, an HDE holder may submit an  
800 HDE supplement requesting that FDA modify the ADN based upon additional information, and  
801 FDA may modify the number.<sup>84</sup>

802 As required under 21 CFR 814.126(b)(1)(iii), the HDE holder (applicant) is responsible for  
803 monitoring how many devices are shipped or sold each year, and if that number exceeds 8,000,  
804 to provide an explanation and estimate to FDA of how the device is being used by patients.  
805 Similarly, the HDE holder is responsible for monitoring when the number of devices shipped or  
806 sold in a year exceeds the ADN, when the HUDs are approved by FDA to make a profit.<sup>85</sup> An  
807 IRB or appropriate local committee is not responsible for monitoring the number of uses per year  
808 of the HUD.

809 If the HDE holder ships multiple sizes of a device to help ensure that one of the devices is the  
810 appropriate size for the patient(s) when used, it would not be necessary to count all of these  
811 devices toward the ADN tally if the additional sizes of the devices (that did not properly fit the  
812 patient(s)) are returned to the HDE holder. Unused devices should be returned to the HDE holder  
813 to appropriately account for them. The HDE holder should document in its periodic report how  
814 many devices are returned to the HDE holder if multiple sizes are shipped. Additionally, HDE  
815 holders should keep in mind that if they distribute devices in excess of the ADN, they will not be  
816 able to make a profit on those devices.

817 HDE holders assigned an ADN must immediately notify the Agency if the number of devices  
818 distributed in a calendar year exceeds the ADN.<sup>86</sup> FDA interprets this statutory requirement to  
819 mean that HDE holders must immediately notify the Agency by submitting an HDE report  
820 whenever the number of HUDs shipped or sold in a calendar year, however the HUD is used,  
821 exceeds the ADN. The statutory notification requirement is generally consistent with the  
822 reporting requirement in 21 CFR 814.126(b)(1)(iii) concerning the number of devices shipped or  
823 sold regardless of their ultimate use (even if outside their approved indications). However, the  
824 statutory provision requires immediate notification when the number shipped or sold in a

---

<sup>84</sup> See section 520(m)(6)(C) of the FD&C Act.

<sup>85</sup> See section 520(m)(6)(A)(iii) of the FD&C Act.

<sup>86</sup> See section 520(m)(6)(A)(iii) of the FD&C Act.

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

825 calendar year exceeds the ADN, whereas the current HDE regulations require periodic reports on  
826 a timeframe specified in the HDE approval order.

827 Once this notification occurs, or once FDA discovers through an inspection that the ADN has  
828 been exceeded, then the sales of the HUD for the remainder of the year are subject to the general  
829 prohibition on profit (unless FDA approves an ADN modification request in an HDE  
830 supplement), and the amount charged for the device must not exceed the cost of research and  
831 development, fabrication, and distribution of the device.<sup>87</sup>

832 In those cases in which a device is approved for a certain indication under an HDE application  
833 and is approved for a different indication under a PMA or De Novo request, sales or shipments  
834 of the device pursuant to the PMA or the De Novo request are not subject to the ADN reporting  
835 requirement. The ADN relates only to those devices that are marketed under an HDE. Therefore,  
836 the manufacturer is required to notify FDA only when sales or shipments pursuant to the HDE  
837 exceed the ADN. If a manufacturer must report the number of sales or shipments of a device  
838 approved for certain indications under a PMA, the manufacturer would be responsible for  
839 separately reporting sales or shipments of devices marketed for different indications under an  
840 HDE per 21 CFR 814.126(b)(1)(iii).

### 841 **C. Information to Patients**

842 Neither the FD&C Act nor FDA regulations require informed consent from patients who are  
843 treated or diagnosed with an HDE-approved HUD in the course of their clinical care. An IRB or  
844 appropriate local committee may, however, choose to require that patients receive certain  
845 information about the HUD when the committee approves use of the HUD for clinical care at a  
846 facility. If a committee requires that patients receive a written document prior to use of the HUD  
847 in clinical care, the document should include much of the information found in the HDE patient  
848 labeling. If no patient information packet is available, the HDE holder may consider including  
849 the following in any written information provided to patients: an explanation that the HUD is  
850 designed to diagnose or treat the disease or condition described in the HDE labeling and that no  
851 comparable device is available to treat the disease or condition; a description of any ancillary  
852 procedures associated with the use of the HUD; a description of the use of the HUD; all known  
853 risks or discomforts; and an explanation of the postulated mechanism of action of the HUD in  
854 relation to the disease or condition. The IRB or appropriate local committee may decide to  
855 require inclusion of this or other information explicitly as part of a written document provided to  
856 patients.

857 The labeling for a HUD approved under an HDE, including any labeling provided to patients,  
858 must be truthful and non-misleading.<sup>88</sup> The device labeling must also include the following  
859 statement clarifying that effectiveness has not been demonstrated: “Humanitarian Device.  
860 Authorized by Federal law for use in the [treatment or diagnosis] of [specify disease or

---

<sup>87</sup> See section 520(m)(6)(D) of the FD&C Act.

<sup>88</sup> See section 502(a) of the FD&C Act, 21 U.S.C. 352(a).

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

861 condition]. The effectiveness of this device for this use has not been demonstrated.”<sup>89</sup> Additional  
862 labeling requirements appear under 21 CFR 814.20(b)(10).

863 **D. HDEs and Pediatric Patients**

864 As discussed above, under section 520(m)(6)(A)(i)(I) of the FD&C Act, a HUD is eligible to be  
865 sold for profit if, among other things, the device is “intended for the treatment or diagnosis of a  
866 disease or condition that occurs in pediatric patients or in a pediatric subpopulation, and such  
867 device is labeled for use in pediatric patients or in a pediatric subpopulation in which the disease  
868 or condition occurs.” This provision permits HDE holders to receive a profit from the sale of  
869 HUDs that are indicated and labeled for pediatric use, subject to the limit of the ADN.

870 HUDs marketed under an HDE may be indicated and labeled for pediatric use only or for use in  
871 both pediatric and adult patients. Devices that are intended to treat both a pediatric population  
872 and an adult population may be included in a single HDE application, but the indications for use  
873 should specify use in pediatric patients, or pediatric subpopulation(s), as well as use in adults. In  
874 some cases, the probable benefit-risk profile for devices intended for use in a pediatric  
875 population, or in a pediatric subpopulation, may differ from its profile when intended for use in  
876 an adult population. Therefore, we recommend that HDE applications for devices intended for  
877 use in pediatric populations and in adult populations include data supporting the use in both  
878 pediatric and adult populations or an appropriate rationale specifically addressing how the data  
879 provided for one population (e.g., adults) are sufficient to support approval of an HDE  
880 application with indications for use in both populations. For more information about  
881 extrapolating data, refer to the FDA guidance, “[Leveraging Existing Clinical Data for](#)  
882 [Extrapolation to Pediatric Uses of Medical Devices](#).”<sup>90</sup>

883 As defined in section 520(m)(6)(E)(i) of the FD&C Act, pediatric patients for purposes of  
884 section 520(m) of the FD&C Act are patients who are 21 years of age or younger (i.e., up to, but  
885 not including, the 22<sup>nd</sup> birthday) at the time of the diagnosis or treatment.<sup>91</sup> As defined by section  
886 520(m)(6)(E)(ii) of the FD&C Act, “pediatric subpopulation” means one of the following  
887 populations: neonates, infants, children, or adolescents. Additional information about the  
888 definition of pediatric patients and pediatric use as it relates to medical devices can be found in  
889 the FDA guidance, “[Premarket Assessment of Pediatric Medical Devices](#).”<sup>92</sup>

890 HUDs that are approved and labeled for pediatric patients or in a pediatric subpopulation as  
891 described in section 520(m)(6)(A)(i)(I) of the FD&C Act are required, under section 520(m)(8)

---

<sup>89</sup> 21 CFR 814.104(b)(4)(ii).

<sup>90</sup> Available at <https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Guidance/Documents/UCM444591>.

<sup>91</sup> See also 21 CFR 814.3(s).

<sup>92</sup> Available at <https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/Guidance/Documents/ucm089742.pdf>.

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

892 of the FD&C Act, to be reviewed annually by FDA’s PAC.<sup>93</sup> The PAC annually reviews these  
893 HUDs to ensure that the HDE remains appropriate for the pediatric populations for which it was  
894 approved. The PAC also conducts periodic review of adverse events for these devices.<sup>94</sup>

895 **E. Review and Approval of the Use of HUDs in**  
896 **Clinical Care**

897 As summarized above, an IRB or appropriate local committee must approve the use of a HUD at  
898 a given facility before it can be used at that facility.<sup>95</sup> Therefore, a health care provider wishing  
899 to use an HDE-approved HUD to treat or diagnose a patient at a facility should obtain approval  
900 from the facility’s IRB or the appropriate local committee before use of the HUD, except in  
901 certain emergencies where prior approval is not required. See Section VIII.G., “Emergency Use  
902 of HUDs.” In reviewing the use of the HUD in clinical care, the IRB or appropriate local  
903 committee should be cognizant that FDA has made a determination of safety and probable  
904 benefit for use of the HUD only within its approved indication(s).

905 The HDE holder is responsible for ensuring that the HUD is administered in facilities that have  
906 oversight by an IRB constituted and functioning in accordance with 21 CFR part 56.<sup>96</sup> Note that  
907 an IRB’s or the appropriate local committee’s approval for the “use” of a HUD at a facility to  
908 treat or diagnose patients in the course of providing clinical care does not mean that there has  
909 been IRB approval of a clinical investigation involving the HUD.

910 FDA interprets the statutory term “appropriate local committee” to mean a standing committee  
911 for the facility that has expertise and experience in reviewing and making treatment decisions for  
912 clinical care, particularly in applying innovative medical device technologies to clinical care. As  
913 such, a standing committee for the facility that includes physicians with experience in the  
914 treatment of rare diseases or conditions would be considered an appropriate local committee by  
915 the Agency. Depending on the facility and the charters of its committees, examples of an  
916 appropriate local committee may include a peer review committee, a credentialing committee, or  
917 a Quality Care Committee. We recommend that the committee include the chief medical officer  
918 or the departmental chief. In addition, FDA interprets the term “appropriate” to mean that  
919 members of the appropriate local committee are free of financial and other conflicts of interest in  
920 decisions pertaining to the use of the HUD in clinical care or they recuse themselves from such  
921 decisions in which they have a conflict of interest. Merely because a facility has a standing  
922 committee does not mean the committee is appropriate to review use of a HUD in clinical care.

923 The IRB or appropriate local committee is not required to review and approve each individual  
924 use of a HUD, nor is it required to audit medical records of patients who receive a HUD. Rather,

---

<sup>93</sup> For more information on the PAC, see <https://www.fda.gov/AdvisoryCommittees/CommitteesMeetingMaterials/PediatricAdvisoryCommittee/default.htm>.

<sup>94</sup> See section 520(m)(7) of the FD&C Act.

<sup>95</sup> Section 520(m)(4)(B) of the FD&C Act.

<sup>96</sup> See section 520(m)(4)(A) of the FD&C Act and 21 CFR 814.124(a).

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

925 the IRB or appropriate local committee may use its discretion to determine how to approve use  
926 of a HUD, including consideration of providers' qualifications through training and expertise to  
927 use the device.<sup>97</sup> For example, with the input of members with the appropriate expertise in the  
928 clinical area, an IRB or appropriate local committee may specify limitations on the use of the  
929 device based upon one or more measures of disease progression, prior use and failure of any  
930 alternative treatment modalities, reporting requirements to the committee or committee  
931 chairperson, appropriate follow-up precautions and evaluations, or other criteria the committee  
932 determines to be appropriate.

## **1. Process and Considerations for Reviewing the Use of HUDs in Clinical Care**

935 For initial review of a HUD, the IRB or appropriate local committee should perform its review at  
936 a convened meeting of the committee.<sup>98</sup> The IRB or appropriate local committee should have  
937 policies and procedures in place for the receipt and evaluation of the materials necessary for  
938 initial approval and continuing review of the HUD's use at that facility. The policies and  
939 procedures should also specify whether the committee requires a consent document for the use of  
940 the HUD at that facility.

941 FDA recommends that the IRB or appropriate local committee follow the review criteria in 21  
942 CFR 56.111 and elsewhere in part 56, where applicable. For example, the IRB or appropriate  
943 local committee should review the risks to patients that are found in the HDE-approved product  
944 labeling, ensure the risks are minimized, and evaluate whether the risks are reasonable in relation  
945 to the proposed use of the device at the facility. FDA also recommends that the IRB or  
946 appropriate local committee review the following materials, as applicable, during initial review  
947 of a request to use a HUD:

- 948 • A copy of the HDE approval order;
- 949 • A description of the device;
- 950 • The product labeling;
- 951 • The patient information packet that may accompany the HUD;
- 952 • A sample consent form for the use of the HUD in clinical care, if required by the IRB or  
953 appropriate local committee or by facility policy; and
- 954 • A summary of how the physician proposes to use the device, including a description of  
955 any screening procedures, the HUD procedure, and any patient follow-up visits, tests or  
956 procedures.

---

<sup>97</sup> For many HDE-approved HUDs, the HDE holder is required to provide training on the use of the device prior to the health care provider using the device. Such requirements would be specified in the HDE application approval order. See 21 CFR 814.126(a) and 814.82(a).

<sup>98</sup> See 21 CFR 56.108, which describes a convened meeting of an IRB for purposes of reviewing FDA-regulated clinical investigations.

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

957 A list of approved HDE applications may be found at [https://www.fda.gov/MedicalDevices/  
958 ProductsandMedicalProcedures/DeviceApprovalsandClearances/HDEApprovals/default.htm](https://www.fda.gov/MedicalDevices/ProductsandMedicalProcedures/DeviceApprovalsandClearances/HDEApprovals/default.htm).  
959 The approval order, labeling, and patient information may be found by selecting the submission  
960 number of the appropriate HDE application.

961 FDA does not require submission of a protocol to the IRB or appropriate local committee for  
962 review when the committee is evaluating a request to use the HUD in the clinical care of patients  
963 at a facility. However, the IRB, appropriate local committee, or institution may require one under  
964 its own policies and procedures.

965 In addition, FDA does not require committees other than the IRB or appropriate local committee  
966 to approve the use of a HUD. However, the institution may require additional review. For  
967 example, the use of another committee to provide assessments of specific risks posed by the  
968 technology or software compatibility may supplement the IRB or appropriate local committee  
969 review.

970 If a physician wants to use a HUD outside its approved indication(s), FDA recommends that the  
971 physician follow the IRB or appropriate local committee's requirements for use of a HUD at that  
972 facility, which may include separate approval requirements for use outside the approved  
973 indication(s). The IRB or appropriate local committee may also require that the physician obtain  
974 informed consent<sup>99</sup> from the patient and ensure that reasonable patient protection measures are  
975 followed, such as devising schedules to monitor the patient, taking into consideration the  
976 patient's specific needs, and the limited information available about the risks and probable  
977 benefits of the device. The extent of oversight in these circumstances is up to the IRB or  
978 appropriate local committee. As discussed above, MDRs must be submitted to FDA and to the  
979 "IRB of record" (i.e., an IRB approving the use of the HUD at the relevant facility) if the device  
980 may have caused or contributed to death or serious injury and for certain malfunctions. If an  
981 appropriate local committee approved the use of the HUD at the facility, FDA recommends that  
982 MDRs be submitted to that committee.

## **2. Continuing Review of the Use of HUDs in Clinical Care**

983  
984  
985 Under FDA's current regulations, an IRB that reviews a request to use a HUD at a facility is  
986 responsible for initial as well as continuing review of the HUD.<sup>100</sup> When an appropriate local  
987 committee conducts such an initial review instead of an IRB, that appropriate local committee  
988 should also conduct continuing review of the HUD. For continuing review, an IRB may use an  
989 expedited review procedure in which a chairperson or one or more experienced reviewers carries  
990 out the review, similar to the expedited review procedure described at 21 CFR 56.110(b). When

---

<sup>99</sup> As noted above, "informed consent" required by a facility in the context of clinical care does not refer to informed consent subject to the requirements in FDA's regulations.

<sup>100</sup> 21 CFR 814.124(a).

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

991 an IRB conducts the initial review, a facility may decide to utilize an appropriate local committee  
992 to conduct continuing review of the use of the HUD in clinical care.

993 Appropriate local committees may develop their own policies and procedures for continuing  
994 review of a HUD and should determine what type of review procedure is appropriate for each  
995 HUD. An expedited procedure, such as that described under 21 CFR 56.110, may be appropriate  
996 for continuing review because a HUD marketed under an HDE is a legally marketed device, and  
997 its use in clinical care does not constitute “research.” An expedited review does not mean a less-  
998 than-substantive review. The individual(s) conducting an expedited review for use of a HUD at a  
999 facility should thoughtfully consider the risk and benefit information available and any MDRs.

1000 In addition, FDA does not require that the IRB or appropriate local committee serve as a Data  
1001 Monitoring Committee. The IRB or appropriate local committee may, however, ask the HDE  
1002 holder for copies of the safety information submitted to FDA in the periodic reports required by  
1003 21 CFR 814.126(b)(1). In this way, information that could have a bearing on human safety  
1004 would be considered at the time of continuing review.

1005 When an IRB or appropriate local committee is deciding whether to approve the use of a HUD  
1006 for clinical care of patients at a facility, it does not make a Significant Risk/Non-Significant Risk  
1007 (SR/NSR) determination. As noted above, use of a legally marketed HUD within its HDE-  
1008 approved indication at a facility to treat or diagnose patients is not a clinical investigation of a  
1009 device under 21 CFR part 812.

## 1010 **F. Review and Approval for Clinical Testing of HUDs**

1011 Clinical investigation of a HUD under an IDE must be approved and supervised by an IRB.<sup>101</sup>  
1012 Data may be collected in a clinical investigation **for the HDE-approved indication(s)** without  
1013 an IDE. An approved IDE permits a device to be shipped lawfully for the purposes of conducting  
1014 investigations of the device without complying with certain other requirements of the FD&C Act  
1015 that would apply to devices in commercial distribution.<sup>102</sup> As long as the HUD is being studied  
1016 for the indication(s) in its approved labeling, the HUD is not subject to IDE requirements  
1017 because the HUD is a legally marketed device and therefore can be lawfully shipped without an  
1018 IDE. However, regardless of the applicability of the IDE regulation at 21 CFR part 812, other  
1019 FDA regulatory requirements may still apply, including among others, requirements for IRB

---

<sup>101</sup> See 21 CFR 56.103, 812.2(b)(ii), and 812.42.

<sup>102</sup> See 21 CFR 812.1.

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

1020 review and approval, financial disclosure, informed consent<sup>103</sup> and, if applicable, additional  
1021 safeguards for children.<sup>104</sup>

1022 If the IRB receives a request to review an investigation to determine safety or effectiveness of  
1023 the HUD for a different indication than the HDE-approved indication(s), then the IRB should be  
1024 aware that this type of clinical investigation is subject to the IDE regulations at 21 CFR part 812.  
1025 If the device is a SR device, the sponsor of the investigation must submit an IDE application to  
1026 FDA and obtain FDA approval of that application before starting the clinical investigation.<sup>105</sup> A  
1027 physician who wants to study a HUD may be the sponsor, investigator, or both for the study. In  
1028 sum, the investigational use of a HUD under these circumstances must be conducted in  
1029 accordance with 21 CFR parts 812, 50, 54, and 56.<sup>106</sup>

#### 1030 Significant Risk/Non-Significant Risk Determinations

1031 An IRB does not have to make a SR/NSR determination when it receives a request to review a  
1032 clinical investigation of a HUD (e.g., collection of safety and effectiveness data) when that  
1033 clinical investigation concerns the HDE-approved indication(s) only. As noted above, FDA does  
1034 not consider such investigations to require an IDE under 21 CFR part 812.

1035 For an investigation of the HUD for indications other than the HDE-approved indication(s), the  
1036 IRB would need to make a SR/NSR determination if that determination has not already been  
1037 made by FDA.<sup>107</sup> In practice, most sponsors have submitted and obtained FDA approval of an  
1038 IDE application before submitting such investigations of HUDs to IRBs for review, so IRBs  
1039 have not needed to make the SR/NSR determination (i.e., FDA had already determined the  
1040 device was a SR device). However, in the event that a sponsor seeks IRB approval for  
1041 investigational use of a HUD for an indication other than its approved indication(s) without first  
1042 obtaining a determination from FDA regarding whether the study is a SR or NSR study, then the  
1043 IRB should make the SR/NSR determination as required in 21 CFR 812.66.

## 1044 **G. Emergency Use of HUDs**

1045 If a physician in an emergency situation determines that IRB or appropriate local committee  
1046 approval for the use of the HUD at the facility cannot be obtained in time to prevent serious harm  
1047 or death to a patient, a HUD may be used without prior approval. In this situation, the HDE

---

<sup>103</sup> Specific requirements for obtaining informed consent from human subjects apply to FDA-regulated clinical investigations. See 21 CFR part 50, subpart B. Note that, in some cases, facilities may have specific requirements for obtaining informed consent for the use of the HDE-approved HUD in the routine clinical care of patients, but these would not be FDA regulatory requirements.

<sup>104</sup> See 21 CFR part 56 for IRB requirements; see 21 CFR part 54 for requirements for financial disclosure by clinical investigators; and see 21 CFR part 50 for requirements for the protection of human subjects, including additional safeguards for children.

<sup>105</sup> 21 CFR 812.20(a).

<sup>106</sup> Note that 45 CFR part 46 may be applicable to research involving HUDs under certain circumstances. The applicability of those regulations is outside the scope of this draft guidance.

<sup>107</sup> See 21 CFR 812.66.

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

1048 holder may ship the HUD, based on the physician’s certification of the emergent need and  
1049 representation that the physician will follow the requirements of 21 CFR 814.124(a) regarding  
1050 reporting. The physician must provide notification of the use to the chairperson of the IRB or  
1051 appropriate local committee, and the notification must include the identification of the patient  
1052 involved, the date of the use, and the reason for the use.<sup>108</sup> FDA regulations require that  
1053 physicians provide such notification to the chairperson of an IRB in writing within 5 days of the  
1054 emergency use of the device. For facilities at which an appropriate local committee reviews the  
1055 use of HUDs instead of an IRB, FDA recommends that physicians also provide the required  
1056 notification of the emergency use in writing and within 5 days.

1057 FDA further recommends that the physician submit a follow-up report on the patient’s condition  
1058 to the HDE holder. The HDE holder is required under 21 CFR 814.126(b) to submit annual  
1059 reports, including the applicant’s clinical experience with the device and the number of devices  
1060 shipped or sold.

DRAFT

---

<sup>108</sup> See section 520(m)(4) of the FD&C Act.

**Contains Nonbinding Recommendations**

**Draft – Not for Implementation**

**Appendix A – Checklist for Filing Review for HDEs**

**(should be completed within 30 days of DCC receipt)**

1061 **HDE Number:** \_\_\_\_\_ **Date Received:** \_\_\_\_\_

1062

1063 **HUD Number (from OOPD):** \_\_\_\_\_

1064

1065 **Device:** \_\_\_\_\_ **Procode:** \_\_\_\_\_

1066 **Company Name/ Address:** \_\_\_\_\_

1067 **Contact Name/Phone Numbers:** \_\_\_\_\_

1068 **FDA Staff Member Name:** \_\_\_\_\_

1069 Within 15 calendar days of receipt of the HDE application, FDA staff should answer the preliminary questions  
 1070 below, which are used as an initial screening of the HDE application. Depending upon the answers to these  
 1071 preliminary questions, the remainder of the filing review may or may not be necessary. If the responses to the  
 1072 preliminary questions and subsequent consultation with FDA staff identified below indicate that the HDE filing  
 1073 review should not continue, the FDA staff member or the CBER regulatory project manager (RPM) should promptly  
 1074 inform the FDA team (including consulting reviewers and management) and notify the requester using proper  
 1075 administrative procedures.

<b>Preliminary Questions</b>		
<b>Answers in the shaded blocks indicate consultation with an identified Center advisor is needed.</b>	<b>Yes</b>	<b>No</b>
<p>1. Is the product a device (per 201(h) of the FD&amp;C Act) or a combination product with a device constituent part? If it appears not to be a device or such a combination product, or you are unsure, consult with the CDRH Jurisdictional Officer or CBER Product Jurisdiction Officer to determine the appropriate action and inform management. <i>Provide summary of Jurisdictional Officer’s/Liaison’s determination.</i></p> <p>If the product does not appear to be a device or a combination product with a device constituent part, mark “No.”</p> <p><b>NOTE:</b> If the product is a combination product with a device constituent part, it may not be appropriate for review under an HDE. If the product is a combination product, consult with the CDRH Jurisdictional Officer (<a href="mailto:cdhrproductjurisdiction@fda.hhs.gov">cdhrproductjurisdiction@fda.hhs.gov</a>) or CBER Product Jurisdiction Officer and inform management.</p>		
<p>2. Is there a copy of, or reference to the determination made by the Office of Orphan Product Development that the device qualifies as a HUD? [814.104(b)(1)]</p> <p>If there is no copy of, or reference to the HUD determination, mark “No.”</p>		
<p>3. If a Request for Designation (RFD) was submitted for the device and assigned to your center, identify the RFD # and confirm the following:</p> <ul style="list-style-type: none"> <li>• Is the device the same (e.g., design, formulation) as that presented in the RFD submission?</li> <li>• Are the indications for use for the device identified in the HDE the same as those identified in the RFD submission?</li> </ul> <p>If you believe the product or the indications presented in the HDE have changed from the RFD, or you are unsure, consult with the CDRH Jurisdictional Officer (<a href="mailto:cdhrproductjurisdiction@fda.hhs.gov">cdhrproductjurisdiction@fda.hhs.gov</a>) or CBER Product Jurisdiction Officer to determine the appropriate action and inform management. <i>Provide summary of Jurisdictional Officer’s/Liaison’s determination.</i></p> <p>If the answer to either question above is no, mark “No.”</p>		

1076 **If the answer to 1 appears to be “No,” then stop review of the HDE and issue the “Original Jurisdictional Product” letter.**

1077 **If the answer to 3 appears to be “No,” then stop the review and contact the CDRH Jurisdictional Officer or CBER**  
 1078 **Product Jurisdiction Officer.**

**Contains Nonbinding Recommendations**

**Draft – Not for Implementation**

<p>4. Is the device eligible for HDE ?</p> <p><b>NOTE:</b> If the device does not appear to be eligible for review through the HDE program because there is a comparable device available (e.g., a predicate device exists, a De Novo request has been granted for a similar device, or an approved PMA exists for a similar device), you should consult with the appropriate CDRH or CBER staff during the filing review.</p> <p>If you believe an application is for a device that is eligible for review through the HDE program and an exemption from the effectiveness provisions, you should (1) complete the 510(k) decision tree to document why the device would be found NSE (<i>attach copy</i>) and (2) obtain concurrence from the appropriate CDRH or CBER staff prior to the filing the original HDE.</p>		
<p>5. Is the applicant the subject of an Application Integrity Policy (AIP)? If “Yes”, consult with the appropriate CDRH Office or CBER Office of Compliance and Biologics Quality/Division of Inspections and Surveillance/Bioresearch Monitoring Branch (OCBQ/DIS/BMB) to determine the appropriate action. Check on web at <a href="https://www.fda.gov/ICECI/EnforcementActions/ApplicationIntegrityPolicy/ucm134453.htm">https://www.fda.gov/ICECI/EnforcementActions/ApplicationIntegrityPolicy/ucm134453.htm</a></p>		

1079 **If the answer to 4 is “No”, the FDA reviewer should consult management and other Center resources to determine the**  
 1080 **appropriate action.**

1081 **If the answer to 5 is “Yes,” then contact CDRH/OC/DBM – BIMO or CBER/OCBQ/DIS/BMB, provide a summary of the**  
 1082 **discussion, and indicate recommendation/action.**

<b>Inventory of Organizational and Administrative Elements                      (Requirements per 21 CFR 814.112 unless otherwise indicated)</b>				
<b>Check “Yes” if item is present, “N/A” if it is not needed and “Not Present” if it is not included but needed.</b>				
	<ul style="list-style-type: none"> <li>• Any “Not Present” answer may result in a “Refuse to File” decision.</li> <li>• Each element on the checklist should be addressed within the application. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (“Yes”). An assessment of the rationale will be considered during the review of the application.</li> </ul>	Present		Not Present (No)
		Yes	N/A	
A.	HDE Content			
	1. Are all required sections in English or accompanied with an English translation?	<input type="checkbox"/>		<input type="checkbox"/>
	2. Is there a table of contents? [814.104(b)(4) and 814.20(b)(2)]	<input type="checkbox"/>		<input type="checkbox"/>
	3. HDE / HUD Information			
	a. Is there an explanation of why the device would not be available unless an HDE was granted? [814.104(b)(2)]	<input type="checkbox"/>		<input type="checkbox"/>
	b. Is there a statement that no other comparable device, other than another approved HUD under an HDE or a device under an approved IDE, is available to treat or diagnose the disease or condition? [814.104(b)(2)]	<input type="checkbox"/>		<input type="checkbox"/>
	c. Is there a discussion of the risks and benefits of currently available devices or alternative forms of treatment? [814.104(b)(2)]	<input type="checkbox"/>		<input type="checkbox"/>
	d. Is there an explanation of why the probable benefit to health from the use of the device outweighs the risk of injury or illness from its use, taking into account the probable risks and benefits of currently available devices or alternative forms of treatment? [814.104(b)(3)]	<input type="checkbox"/>		<input type="checkbox"/>
	e. Has the amount to be charged for the device been provided, and if greater than \$250.00, is a report provided verifying that the amount charged does not exceed the costs of the device’s research, development, fabrication, and distribution? [814.104(b)(5)]	<input type="checkbox"/>		<input type="checkbox"/>

**Contains Nonbinding Recommendations**

**Draft – Not for Implementation**

<b>Inventory of Organizational and Administrative Elements (Requirements per 21 CFR 814.112 unless otherwise indicated)</b>										
<b>Check “Yes” if item is present, “N/A” if it is not needed and “Not Present” if it is not included but needed.</b>										
		<ul style="list-style-type: none"> <li>Any “Not Present” answer may result in a “Refuse to File” decision.</li> <li>Each element on the checklist should be addressed within the application. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (“Yes”). An assessment of the rationale will be considered during the review of the application.</li> </ul>		<table border="1"> <tr> <th colspan="2">Present</th> <th rowspan="2">Not Present (No)</th> </tr> <tr> <th>Yes</th> <th>N/A</th> </tr> </table>		Present		Not Present (No)	Yes	N/A
Present		Not Present (No)								
Yes	N/A									
4.	Is a bibliography provided? [814.104(b)(4) and 814.20(b)(8)(i)]		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>					
	a.	Have copies of key articles been provided and are English translations included, if appropriate? Check “N/A” if applicant includes a statement that upon searching they found no literature related to their device	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>					
5.	If a device sample has been requested by FDA, has it been provided or if impractical to submit, has the applicant offered alternatives to allow FDA staff to view or access the device? [814.104(b)(4) and 814.20(b)(9)]		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>					
6.	Is there a summary of the contents of the HDE? [814.104(b)(4) and 814.20(b)(3)]		<input type="checkbox"/>		<input type="checkbox"/>					
7.	Device Characteristics									
	a.	Is a description of the device included? [814.104(b)(4) and 814.20(b)(4)]	<input type="checkbox"/>		<input type="checkbox"/>					
		i. Pictorial representations? [814.104(b)(4) and 814.20(b)(4)(i)]	<input type="checkbox"/>		<input type="checkbox"/>					
		ii. Materials specifications? [814.104(b)(4) and 814.20(b)(4)(i)-(ii)]	<input type="checkbox"/>		<input type="checkbox"/>					
		<ul style="list-style-type: none"> <li>If there is a color additive present:                             <ul style="list-style-type: none"> <li>has the color additive been identified by common name and chemical name, and</li> <li>has the amount of each color additive in the formulation by weight percent of the colored component and total amount (e.g., ppm, µg) in the device been provided?</li> </ul> </li> </ul>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>					
	b.	Is a description of the principles of operation of the device (including components) and properties relevant to clinical function present? [814.104(b)(4) and 814.20(b)(4)(iii)-(iv)]	<input type="checkbox"/>		<input type="checkbox"/>					
8.	Is the Device Manufacturing Section included (see the FDA guidance, “Quality System Information for Certain Premarket Application Reviews,” <a href="https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm070899.pdf">https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm070899.pdf</a> ) [814.104(b)(4) and 814.20(b)(4)(v)]		<input type="checkbox"/>		<input type="checkbox"/>					
	a.	Has a description of the methods, facilities, and controls used in the manufacture, processing, packing, storage, and installation of the device been provided?	<input type="checkbox"/>		<input type="checkbox"/>					
9.	The application includes a summary and full study report* for each nonclinical study provided? [814.104(b)(4) and 814.20(b)(6)(i)]		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>					
	<p>Note: the applicant can reference data located in other applications. Check “Yes” if nonclinical data is not provided in the current application but found in another application. State where the data were provided (e.g., modular application, master file).</p> <p>*Full study report includes objective of the test, description of test methods and procedures, study endpoint(s), pre-defined pass/fail criteria, results summary, and discussion of conclusions. In the event that an applicant is appropriately declaring conformity with a voluntary consensus standard to meet applicable requirements, full test</p>									

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

<b>Inventory of Organizational and Administrative Elements (Requirements per 21 CFR 814.112 unless otherwise indicated)</b>					
<b>Check “Yes” if item is present, “N/A” if it is not needed and “Not Present” if it is not included but needed.</b>					
		<ul style="list-style-type: none"> <li>Any “Not Present” answer may result in a “Refuse to File” decision.</li> <li>Each element on the checklist should be addressed within the application. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (“Yes”). An assessment of the rationale will be considered during the review of the application.</li> </ul>	Present		Not Present (No)
			Yes	N/A	
		reports may not be necessary with respect to those requirements. Refer to 13(a).			
	a.	Sterilization	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	b.	Biological/Microbiological	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	c.	Immunological	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	d.	Toxicological/Biocompatibility	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	e.	Engineering (stress, wear, etc.)	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	f.	Chemistry/Analytical (typically for IVDs)	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	g.	Shelf Life	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	h.	Animal Studies	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	i.	Other Essential Laboratory Testing	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	10.	Is a summary of clinical experience and investigation(s) and results provided? [814.104(b)(4)(i) as applicable]	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	a.	Are the final versions of the clinical protocols included? (If performed under IDE, these should be the final FDA-approved versions of the clinical protocols, incorporating any Notices of Changes.)	<input type="checkbox"/>		<input type="checkbox"/>
	b.	Is a description of study population demographics provided?	<input type="checkbox"/>		<input type="checkbox"/>
	c.	Is a description of adverse events (e.g., adverse reactions, complaints, discontinuations, failures, replacements) provided?	<input type="checkbox"/>		<input type="checkbox"/>
	d.	Have report forms for patients who died or who did not complete the investigation been provided (i.e., to resolve potential bias)?  Check “N/A” only if no patients died or were discontinued.	<input type="checkbox"/>		<input type="checkbox"/>
	11.	Are statistical analyses of the clinical investigations provided, if appropriate? [814.104(b)(4)(i)]	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	a.	Are the results of all analyses identified in the protocol provided?	<input type="checkbox"/>		<input type="checkbox"/>
	12.	Has appropriate draft labeling been submitted? [814.104(b)(4) and 814.20(b)(10)]			
	a.	Physician Labeling	<input type="checkbox"/>		<input type="checkbox"/>
	i.	Are indications for use included?	<input type="checkbox"/>		<input type="checkbox"/>
	ii.	Are contraindications, warnings, and precautions included?	<input type="checkbox"/>		<input type="checkbox"/>
	iii.	Are instructions for use included?	<input type="checkbox"/>		<input type="checkbox"/>
	iv.	Does the labeling include the statement: “Humanitarian Device. Authorized by Federal law for use in the [treatment or diagnosis] of [specify disease or condition]. The effectiveness of this device for this use has not been	<input type="checkbox"/>		<input type="checkbox"/>

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

<b>Inventory of Organizational and Administrative Elements (Requirements per 21 CFR 814.112 unless otherwise indicated)</b>							
Check “Yes” if item is present, “N/A” if it is not needed and “Not Present” if it is not included but needed.							
					Present		Not Present (No)
					Yes	N/A	
				demonstrated” [814.104(b)(4)(ii)]			
		b.		Patient Labeling Check  Check “N/A” only if the relevant lead Center has previously indicated that patient labeling is not necessary.	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		c.		Technical/Operators Manual, if applicable	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	13.			Statements/Certifications/Declarations of Conformity [814.104(b)(4), 814.20(b)(5), and 814.20(b)(12)]			
		a.		Has the applicant provided documentation to establish conformance with applicable performance standards and/or voluntary consensus standards?  Check “N/A” only if no standards are used.	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		b.		Has the applicant provided documentation to establish that it has followed the recommendations in applicable FDA guidance or otherwise met applicable statutory or regulatory criteria? Check “N/A” only if no guidance is used.	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		c.		Investigator Financial Disclosure For additional information refer to the guidance document “Guidance for Industry – Financial Disclosure by Clinical Investigators” ( <a href="https://www.fda.gov/RegulatoryInformation/Guidances/UCM341008">https://www.fda.gov/RegulatoryInformation/Guidances/UCM341008</a> )  As required by 21 CFR Part 54, has the applicant submitted for each clinical investigator either: 1. A signed and dated Certification Form (3454) or 2. A signed and dated Disclosure Form (3455)  Note: the signature should be from a responsible corporate official or representative of the applicant.	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		i.		For a Certification Form (3454): Is the required list of all investigators and subinvestigators attached to the Form?	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		ii.		If box 3 of Form 3454 is checked, does the Form include an attachment with the reason(s) why financial disclosure information could not be obtained?	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		iii.		For a Disclosure Form (3455): Does the application provide details of the financial arrangements and interests of the investigator(s) or subinvestigator(s), along with a description of any steps taken to minimize potential bias?	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		d.		Environmental Assessment under 21 CFR 25.20(n) [814.104(b)(4) and	<input type="checkbox"/>		<input type="checkbox"/>

**Contains Nonbinding Recommendations**

**Draft – Not for Implementation**

<b>Inventory of Organizational and Administrative Elements (Requirements per 21 CFR 814.112 unless otherwise indicated)</b>						
<b>Check “Yes” if item is present, “N/A” if it is not needed and “Not Present” if it is not included but needed.</b>						
<ul style="list-style-type: none"> <li>Any “Not Present” answer may result in a “Refuse to File” decision.</li> <li>Each element on the checklist should be addressed within the application. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (“Yes”). An assessment of the rationale will be considered during the review of the application.</li> </ul>				Present		Not Present (No)
				Yes	N/A	
			814.20(b)(11)]			
		i.	If claiming a categorical exclusion, information to justify the exclusion, OR	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		ii.	An environmental assessment ( <u>ONLY</u> required for devices that present new environmental concerns)	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		e.	Did the application include a completed FORM FDA 3674, <i>Certification with Requirements of ClinicalTrials.gov Data Bank?</i> (42 U.S.C. 282(j)(5)(B) and 42 CFR part 11)  Note: Enter the NCT number(s) in the Center Tracking System (CTS)	<input type="checkbox"/>		<input type="checkbox"/>
			Data from FORM FDA 3674 (mark “Yes” for the applicable one):			
		i.	No clinical trials referenced in application.	<input type="checkbox"/>	<input type="checkbox"/>	
		ii.	Requirements are not applicable to referenced clinical trials.	<input type="checkbox"/>	<input type="checkbox"/>	
		iii.	Requirements are applicable and have been met.	<input type="checkbox"/>	<input type="checkbox"/>	
	14.		Pediatric Use - Per 515A(a)(2) of the FD&C Act, did the application include, if readily available: [814.104(b)(6) and 814.20(b)(13)]			
		a.	A description of any pediatric subpopulations that suffer from the disease or condition that the device is intended to treat, diagnose, or cure, or statement that no pediatric subpopulation exists for the disease or condition for which the device is intended. This statement does not mean the device is indicated for treating pediatric patients. For additional information refer to the guidance document “Providing Information about Pediatric Uses of Medical Devices - Guidance for Industry and Food and Drug Administration Staff” at <a href="https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM339465.pdf">https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM339465.pdf</a> .	<input type="checkbox"/>		<input type="checkbox"/>
		b.	The number of affected pediatric patients.	<input type="checkbox"/>		<input type="checkbox"/>
B.	Issues Identified by FDA Prior to receipt of the HDE Application - history of the applicant with this device					
	1.	Does the applicant list prior applications or state that there were no prior applications? (may be located in CDRH Coversheet Form FDA 3514, Section F)  If the applicant lists prior applications, address the applicable questions below:		<input type="checkbox"/>		<input type="checkbox"/>
		a.	510(k) # _____	<input type="checkbox"/>	<input type="checkbox"/>	
		i.	If this device has been the subject of an NSE decision, have the data presented in the HDE taken into account any concerns related to safety or probable benefit that were previously communicated during the review of the prior 510(k) or through 510(k) correspondence?	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

<b>Inventory of Organizational and Administrative Elements (Requirements per 21 CFR 814.112 unless otherwise indicated)</b>						
<b>Check “Yes” if item is present, “N/A” if it is not needed and “Not Present” if it is not included but needed.</b>						
		<ul style="list-style-type: none"> <li>Any “Not Present” answer may result in a “Refuse to File” decision.</li> <li>Each element on the checklist should be addressed within the application. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (“Yes”). An assessment of the rationale will be considered during the review of the application.</li> </ul>		<b>Present</b>		<b>Not Present (No)</b>
				<b>Yes</b>	<b>N/A</b>	
	b.	IDE # _____		<input type="checkbox"/>	<input type="checkbox"/>	
	i.	Have the data presented in the HDE taken into account any safety or probable benefit concerns (e.g., “future considerations”) previously communicated during the review of prior IDE(s) or through IDE correspondence?		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	c.	PMA # _____		<input type="checkbox"/>	<input type="checkbox"/>	
	i.	If a previously submitted PMA for this device has been withdrawn or denied, does the current HDE application take into account any issues related to safety or probable benefit raised during review of the prior PMA(s) or through PMA correspondence?		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	d.	HDE # _____		<input type="checkbox"/>	<input type="checkbox"/>	
	i.	If a previously submitted HDE application for this device has been withdrawn or denied, does the current HDE application take into account any issues related to safety or probable benefit raised during review of the prior HDE application or through HDE correspondence?		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	e.	Modular HDE # _____		<input type="checkbox"/>	<input type="checkbox"/>	
	i.	If “Yes”, how many modules submitted? _____ How many modules were closed? _____				
	ii.	If there are modules that are on hold, does the HDE address outstanding deficiencies?		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	2.	Does the applicant list Pre-Submission(s) regarding the device or this application in which FDA feedback regarding data or information related to safety and/or probable benefit in the HDE was provided by email or during a meeting (in person or by phone), or state that there were no prior Pre-Submission interactions with the FDA regarding this application?  If the applicant lists Pre-Submissions, address the applicable questions below:		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	a.	Pre-Submission # _____ Meeting date(s), if applicable _____				
	b.	Copy of minutes from each meeting or other written feedback?		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	c.	Were all FDA concerns or action items previously presented to the applicant in the Pre-Submission minutes or feedback addressed in the HDE or has the applicant provided a detailed scientific or clinical justification for an alternative approach?		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

<b>Filing Decision Questions</b>			
A “No” answer will typically result in a Not-Filed decision.			
		Yes	No
Decision 1	<p>Is the HDE complete?</p> <p>If, on its face, the HDE is missing one or more required elements (identified above), such that the application is not sufficiently complete to permit substantive review, answer “No.”</p>	<input type="checkbox"/>	<input type="checkbox"/>
Decision 2	<p>From only an administrative review, does the HDE include information that appears to constitute valid scientific evidence?</p> <p>Only answer “No” if it is clear that the HDE is supported solely by information that 21 CFR 860.7 identifies as not constituting valid scientific evidence:</p> <ul style="list-style-type: none"> <li>• isolated case reports</li> <li>• random experience</li> <li>• reports lacking sufficient details to permit scientific evaluation</li> <li>• unsubstantiated opinions</li> </ul> <p>Comments:</p>	<input type="checkbox"/>	<input type="checkbox"/>
Decision 3	<p>Does the HDE address the key nonclinical and clinical issues identified by FDA prior to submission of the HDE application?</p> <p><b>OR</b></p> <p>Has the applicant provided a detailed scientific or clinical justification for the alternate approach?</p> <p>Section B of the checklist outlines questions intended to identify when the FDA has previously provided specific guidance to the applicant about the content of the HDE application through one or more mechanisms, such as a prior HDE or PMA application, a prior “Not Substantially Equivalent” decision on a 510(k), Investigational Device Exemption (IDE) letters, feedback through the Q-submission Program, a Determination or Agreement meeting(s), or other substantive communication with FDA, or through a published guidance document. If such information has been communicated to the applicant through one or more of these mechanisms, and the HDE application addresses each of the key nonclinical and clinical issues identified by FDA, the answer to the above question is “Yes.” Furthermore, if some of these key issues previously identified by FDA are not addressed, but the HDE application contains a scientific or clinical justification for the omission or deviation, the answer to the above question is “Yes.” These cases do not preclude the responsible review Division from accepting the HDE application.</p> <p>In this context, the term “key issues” is meant to refer to issues that are central to FDA’s review of the device’s safety and probable benefit under sections 515 and 520(m) of the FD&amp;C Act. Examples of key issues include: need for long-term nonclinical studies (e.g., biocompatibility, carcinogenicity, or other animal studies), and certain clinical study parameters (e.g., sample size, patient population, study design, and endpoints). These key issues are typically device-specific. As a result, the decision of FDA to “Refuse to File” an HDE application based on this criterion can only be made after carefully considering the following questions:</p> <p><i>Are the types of necessary nonclinical and clinical studies well-known in the scientific and</i></p>	<input type="checkbox"/>	<input type="checkbox"/>

***Contains Nonbinding Recommendations***

***Draft – Not for Implementation***

<b>Filing Decision Questions</b>			
<b>A “No” answer will typically result in a Not-Filed decision.</b>			
		<b>Yes</b>	<b>No</b>
	<p><i>medical communities for the particular device?</i></p> <p>For an “established” device type, the types of nonclinical and clinical studies that we would expect in a PMA are likely to be well-known both within FDA and in the scientific and medical communities and, as such, are often included as part of an FDA guidance document and/or consensus standard. You should bear in mind that, for HDEs, the device may not be of an established device type.</p> <p><i>Were the issues conveyed to the applicant as part of a documented regulatory process?</i> Examples of a documented regulatory process include:</p> <ul style="list-style-type: none"> <li>• interaction through the Q-submission process,</li> <li>• prior PMA or HDE application,</li> <li>• prior “Not Substantially Equivalent” decision on a 510(k),</li> <li>• IDE letters, or</li> <li>• letter(s) issued as a result of Determination or Agreement meetings.</li> </ul> <p><i>Were the issues conveyed to the applicant related to insufficient effectiveness data?</i></p> <p>Devices approved under an HDE application are exempt from the requirement to demonstrate a reasonable assurance of effectiveness. If an issue relates to insufficient effectiveness data, filing the HDE may be appropriate in cases for which accepting a PMA would not.</p> <p>FDA staff should only designate an HDE “Refuse to File” based on a “No” response to “Acceptance Decision 3” in instances where the key issues were identified by FDA staff as part of a documented regulatory process.</p>		

1084    **Decision:** *FDA Staff Recommendation: File* \_\_\_\_ *Not File* \_\_\_\_

1085 **Appendix B – Considerations for the Probable Benefit-Risk**  
1086 **Assessment**

1087 As discussed in Section VI of this guidance, FDA considers the same factors described in FDA’s  
1088 benefit-risk framework for evaluating PMAs or De Novo requests when assessing probable  
1089 benefits and risks for HDE applications. Refer to the FDA guidance document, “Factors to  
1090 Consider when Making Benefit-Risk Determinations in Medical Device Premarket Approval and  
1091 De Novo Classifications,” for a description of those factors. It should be clearly noted, however,  
1092 that probable benefit and probable benefit-risk determinations under an HDE are different from  
1093 those under a PMA or a De Novo request. Please refer to Sections V and VI of the guidance for  
1094 further discussion related to these differences and the probable benefit-risk assessment. The tools  
1095 identified in Appendices B and C are meant to serve complementary roles, and both should be  
1096 completed as part of the probable benefit-risk assessment.

1097 *Instructions:* FDA staff should make their recommendation regarding the probable benefit-risk  
1098 assessment based on the totality of the evidence. The probable benefit-risk assessment is part of  
1099 the decision whether to approve the application, but it does not include an assessment of all  
1100 applicable requirements for approval. An indication from these tools that the probable benefits  
1101 outweigh the risks does not mean that the application satisfies other applicable requirements for  
1102 an HDE application.

1103 The following questions are intended as a sequential method to help weigh various factors as part  
1104 of the probable benefit-risk assessment. As such, the questions are intended to help identify and  
1105 explain which factors and considerations are critical in making a probable-benefit risk  
1106 assessment for a particular device. However, the questions are not intended to suggest that  
1107 considerations other than those listed in the completed worksheet are irrelevant.

1108 Consider questions 1-8 for Column A (the proposed Indications for Use), to help determine if the  
1109 application is approvable for the proposed indications or, if narrowed indications for use are  
1110 appropriate and can be agreed upon with the sponsor, consider questions 1-8 in Column B.  
1111 However, as reflected under question 1, if the evidence does not support a finding of probable  
1112 benefit under the proposed Indications for Use (or narrowed Indications for Use), or evidence  
1113 does not support a finding of probable benefit for the proposed Indications for Use and  
1114 agreement on narrowed Indications for Use is not achievable or applicable, the application would  
1115 not be approvable.

1116 **Assessment of Probable Benefit**

1117 **1. Is there any evidence of clinical benefit?**

1118 *Complete question 1a for therapeutics or invasive/implantable diagnostics; complete question 1b*  
1119 *for other diagnostics.*

1120  
1121 *Note that in lieu of summaries, conclusions, and results of clinical investigations required under 21*  
1122 *CFR 814.20(b)(3)(v)(B), (b)(3)(vi), and (b)(6)(ii), HDE applicants are required to submit summaries,*  
1123 *conclusions, and results of all clinical experience or investigations (whether adverse or*

## Contains Nonbinding Recommendations

### Draft – Not for Implementation

1124 supportive) reasonably obtainable by the applicant that are relevant to an assessment of the risks  
1125 and probable benefits of the device (see 21 CFR 814.104(b)(4)(i)).

1126  
1127 a. Therapeutics or invasive/implantable diagnostics: Is a clinical benefit demonstrated for  
1128 the device for this indication (e.g., from any one or more of the primary and/or secondary  
1129 datasets or from associated real world evidence)? Probable benefit may be considered in  
1130 terms of how a patient feels, functions, survives, or an acceptable surrogate outcome.  
1131 Probable benefit may also be considered in terms of convenience in managing a disease  
1132 or condition. *Select any of the following that demonstrate benefit.*

1133 Indication

A B

- A favorable change in at least 1 clinical assessment which is equal to or greater than seen in the control group
- A favorable change in at least 1 clinical assessment which meets a predetermined performance goal
- A favorable change in at least 1 clinical assessment which meets or surpasses a minimally important clinical difference
- A favorable change in at least 1 clinical assessment which is equal to or greater than seen with other available modalities for the disease or condition
- A favorable change in at least 1 clinical assessment which would be meaningful to patients considering the severity, chronicity, etc., of the condition, taking into consideration patient-reported outcomes and health-related quality of life
- A favorable change in non-clinical data or modeling that is deemed to be predictive of clinical outcomes
- Other(s) [Click here to list other\(s\)](#)

1134  
1135 b. Other diagnostics: Is a clinical benefit demonstrated, based on accurate measurement of  
1136 the diagnostic analyte(s)/biomarker in the indicated population? *Select one or more of the*  
1137 *following clinical indications that demonstrate benefit, if applicable.*

1138 Indication

A B

- Disease susceptibility (likely future occurrence of a disease or condition)
- Screening
- Diagnosis
- Prognosis
- Monitoring
- Treatment selection/modification
- Other(s) [Click here to list other\(s\)](#)

1139  
1140 **Question 1: Is there any evidence of clinical benefit?**

1141 Indication

A B

- YES → Continue to Question 2
- NO → Move one column to the right or, if final column has been reached and you have determined there is no evidence of clinical benefit, do not approve the application.

## Contains Nonbinding Recommendations

### Draft – Not for Implementation

1142  
1143

#### 2. What is the degree of uncertainty for the probable benefits?

1144 a. Recognizing that some degree of uncertainty always exists, select the elements that  
1145 contribute to uncertainty, if applicable, in the data regarding probable clinical benefit.

##### 1146 Indication

A B

- Inconsistent or conflicting results between studies
- Wide confidence intervals surrounding the point estimate(s) and/or odds ratio(s)
- High subject loss-to-follow-up at critical assessment point(s)
- Large amount of missing data at critical assessment time(s) +/- imputation
- Significant number of major protocol deviations
- Impact of confounding interventions
- Inconsistent user experience or user experience not representative of likely real world user
- Non-clinical data or modeling that does not adequately represent the clinical circumstance
- Real World Evidence (RWE) is not relevant or reliable for the purposes of the proposed analysis
- Inspectional findings
- Study results are not generalizable to population under consideration
- Other(s) [Click here to list other\(s\)](#)

1147  
1148  
1149  
1150

b. Diagnostics: Select the performance characteristics that contribute to uncertainty for analytical validation of the device:

##### Indication

A B

- Sensitivity
- Specificity/Interference
- Accuracy
- Precision
- Reproducibility
- Reportable range or status
- Linearity/Recovery
- Matrix, carryover
- Calibrators and/or controls
- Pre-analytical
- Post-analytical
- Other(s) [Click here to list other\(s\)](#)

1151  
1152  
1153

#### **Question 2: What is the degree of uncertainty for the probable benefits?**

##### Indication

A B

- Low → Continue to Question 3; consider suggesting a different kind of marketing application.
- Medium → Continue to Question 3
- High → Continue to Question 3

## Contains Nonbinding Recommendations

### Draft – Not for Implementation

1154

#### 1155 **Comments on the Assessment of Probable Benefit**

##### 1156 **For the Proposed Indications for Use (Column A):**

1157 Click here to enter summary of the Assessment of Probable Benefit for the proposed Indications for Use.

1158 Include a description of your assessment of the extent of probable benefit, considering the type,  
1159 magnitude, and probability of benefit(s); and the duration of effects. Include a description of the impact of  
1160 uncertainty on your Assessment of Probable Benefit.

1161

##### 1162 **For the narrowed Indications for Use (modified indication and/or population) (Column B):**

1163 Click here to enter summary of the Assessment of Probable Benefit for the narrowed Indications for Use.

1164 Include a description of your assessment of the extent of probable benefit, considering the type,  
1165 magnitude, and probability of benefit(s); and the duration of effects. Include a description of the impact of  
1166 uncertainty on your Assessment of Probable Benefit.

## 1167 **Assessment of Risk**

### 1168 **3. Are known/probable risks more than minimal?**

1169 *Select the elements that apply for known/probable risks that are more than minimal.*

#### 1170 Indication

A B

- Adverse events (AEs) or outcomes related to the device itself
- AEs or outcomes related to the use of the device or procedure to use the device
- AEs or outcomes related to anesthesia or sedation to use the device
- AEs or outcomes due to subsequent tests/treatments needed (e.g., radiation from CT scans)
- AEs or outcomes, not seen in the study/data, but probable based on “class effect” or events known to occur with similar technologies
- False positive/false negative/absent result for diagnostics
- Other(s) Click here to list other(s)

1171

### 1172 **Question 3: Are known/probable risks more than minimal?**

#### 1173 Indication

A B

- YES → Continue to Question 4
- NO → Continue to Question 4

1174

### 1175 **4. What is the degree of uncertainty for the risks?**

1176 *Recognizing that some degree of uncertainty always exists, select the elements that contribute to*  
1177 *uncertainty, if applicable, in the data regarding the adverse events/outcomes or risks.*

#### 1178 Indication

A B

- Insufficient patient/subject numbers to detect serious events
- Insufficient duration of follow-up to detect delayed/late events

## Contains Nonbinding Recommendations

### Draft – Not for Implementation

A B

- Lack of data on repeated exposure to the device/use
- Inconsistent or conflicting results between studies
- Proper evaluations not performed as part of the study protocol to adequately detect certain AEs
- Poor or inconsistent adverse event definitions and documentation
- Events likely confounded by, and attributed to, other comorbidities or treatment modalities
- High subject loss-to-follow-up at critical assessment point(s)
- Large amount of missing data at critical assessment time(s) +/- imputation
- Significant number of major protocol deviations
- Inconsistent user experience or user experience not representative of likely real world user
- False positive/false negative/absent/indeterminate result for diagnostics
- Other(s) [Click here to list other\(s\)](#)

1179

1180

#### **Question 4: What is the degree of uncertainty for the risks?**

1181

##### Indication

A B

- Low → Continue to Question 5
- Medium → Continue to Question 5
- High → Continue to Question 5

1182

1183

#### **Comments on the Assessment of Risk**

1184

**If you answered “No” to Question 3 but “High” to Question 4, please explain here.**

1185

1186

#### **For the Proposed Indications for Use (Column A):**

1187

Click here to enter summary of the Assessment of Risk for the proposed Indications for Use. Include a description of your assessment of the extent of probable risk considering the severity, types, number and rates of harmful events associated with use of the device; probability of a harmful event; duration of harmful events; and risk from false-positive or false-negative results for diagnostics. Include a description of the impact of uncertainty on your Assessment of Risk.

1192

1193

#### **For the narrowed Indications for Use (modified indication and/or population) (Column B):**

1194

Click here to enter summary of the Assessment of Risk for the narrowed Indications for Use. Include a description of your assessment of the extent of probable risk, considering the severity, types, number and rates of harmful events associated with use of the device; probability of a harmful event; duration of harmful events; and risk from false-positive or false-negative results for diagnostics. Include a description of the impact of uncertainty on your Assessment of Risk.

1198

1199

## **Assessment of Probable Benefit-Risk**

1200

*Provide a recommendation based on the totality of the evidence. As noted above, the probable benefit-risk assessment is part of the decision regarding whether to approve an HDE application but is not an assessment of all applicable requirements.*

1201

1202

1203

### **5. Do the Probable Benefits outweigh the Risks, considering the assessment of Probable Benefit and Risk and the degree of uncertainty identified above, and taking into account the**

1204

## Contains Nonbinding Recommendations

### Draft – Not for Implementation

1205 **probable benefits and risks of currently available devices or alternative forms of treatment?**

1206

1207 *To approve an HDE application, FDA must find, among other things, that the device will not expose*  
1208 *patients to an unreasonable or significant risk of illness or injury and that the probable benefit to health*  
1209 *from the use of the device outweighs the risk of injury or illness from its use, taking into account the*  
1210 *probable benefits and risks of currently available devices or alternative forms of treatment. Consider how*  
1211 *the probable benefits and risks identified above compare to currently available devices or alternative forms*  
1212 *of treatment and select the elements that apply:*

1213 Indication

A B

- No legally marketed alternative medical product or medical intervention exists or the device offers advantages over other modalities
- The device fills an unmet medical need or niche for more effective treatment or diagnosis of life-threatening or irreversibly debilitating human disease/conditions
- The probable benefit of the device is equivalent to or greater than the probable benefits of other modalities
- The probable risk of the device is no greater or is less than the probable risks of other modalities
- The probable benefit of the device is less than the probable benefits of other modalities
- The probable benefit of the device is less than the probable benefits of other modalities
- The probable risk of the device is greater than the probable risks of other modalities
- Other(s) [Click here to list other\(s\)](#)

1214

1215 **Question 5: Do the Probable Benefits outweigh the Risks, considering the assessment of**  
1216 **Probable Benefit and Risk and the degree of uncertainty identified above, and taking into**  
1217 **account the probable benefits and risks of currently available devices or alternative forms**  
1218 **of treatment?**

1219 Indication

A B

- Yes – The probable benefits outweigh the risks such that, for this device, additional consideration of relevant factors would not change that determination.
- Undetermined – The probable benefits may not outweigh the risks, and further discussion and consideration of relevant factors is appropriate → Move to Question 6

1220

1221 **Comments on Assessment of Probable Benefit-Risk**

1222 **For the Proposed Indications for Use (Column A):**

1223 [Click here to summarize the probable benefit\(s\) that have been demonstrated for the proposed Indications](#)  
1224 [for Use and your assessment of how Probable Benefit\(s\) compare to Risks. Include a description of how](#)  
1225 [available alternative modalities, including their probable benefits and risks, affect your assessment.](#)

1226 [Include a description of how uncertainty regarding Probable Benefit\(s\) and Risks affects your assessment.](#)

1227

1228 **For the narrowed Indications for Use (modified indication and/or population) (Column B):**

1229 [Click here to summarize the probable benefit\(s\) that have been demonstrated for the narrowed Indications](#)  
1230 [for Use and your assessment of how the Probable Benefit\(s\) compare to Risks. Include a description of](#)  
1231 [how available alternative modalities, including their probable benefits and risks, affect your assessment.](#)

1232 [Include a description of how uncertainty regarding Probable Benefit\(s\) and Risks affects your assessment.](#)

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

1233  
1234  
1235  
1236

**6. Do the Probable Benefits outweigh the Risks, when taking into account additional relevant considerations? *Select relevant considerations.***

Indication

A B

- Available patient preference information (PPI) showing patients willingness or unwillingness to accept the probable risks in exchange for the probable benefits
- Available qualitative or quantitative PPI on the relative desirability or acceptability to patients of outcomes or other attributes that differ among alternative health interventions
- Understanding that the device represents novel technology for which the current device technology is different
- Ability to manage the condition and consideration of natural history of disease progression in the absence of the intervention with the device under review
- The adverse events associated with use of the device are reversible
- Type of intervention required to address the harmful event (e.g., medication, surgery)
- Understanding of mechanistic plausibility and/or “class effect” (e.g., familiarity with similar technology)
- Other(s) [Click here to list other\(s\)](#)

1237  
1238  
1239  
1240

**Question 6: Do the Probable Benefits outweigh the Risks, when taking into account additional relevant considerations?**

Indication

A B

- Yes –The probable benefits outweigh the risks such that, for this device, additional consideration of relevant factors would not change that determination.
- Undetermined – The probable benefits may not outweigh the risks, and discussion and consideration of risk mitigation measures is appropriate → Move to Question 7

1241

**Comments on Assessment of Probable Benefit-Risk, taking into account additional relevant considerations**

1242

**For the Proposed Indications for Use (Column A):**

1243

Click here to summarize the probable benefit(s) that have been demonstrated for the proposed Indications for Use and your assessment of how Probable Benefit(s) compare to Risks. Include a description of how available alternative modalities, including their probable benefits and risks, affect your assessment.

1244

Include a description of how uncertainty regarding Probable Benefit(s) and Risks affects your assessment.

1245

Include a description of how patient perspectives affected your assessment.

1246

1247

1248

1249

1250

**For the narrowed Indications for Use (modified indication and/or population) (Column B):**

1251

Click here to summarize the probable benefit(s) that have been demonstrated for the narrowed Indications for Use and your assessment of how Probable Benefit(s) compare to Risks. Include a description of how available alternative modalities, including their probable benefits and risks, affect your assessment.

1252

Include a description of how uncertainty regarding Probable Benefit(s) and Risks affects your assessment.

1253

Include a description of how patient perspectives affected your assessment.

1254

1255

1256

1257

***Contains Nonbinding Recommendations***

***Draft – Not for Implementation***

1258 **7. Can the risks be mitigated, so that Probable Benefits outweigh the Risks? Consider if the**  
1259 **Probable Benefits outweigh the Risks if risk mitigation strategies are incorporated to lower**  
1260 **the probability of a harmful event occurring and improve the probable benefit-risk profile**  
1261 **of the device. Select relevant considerations:**

1262 Indication

A B

- Additional descriptions of known and probable benefits and risks in physician and patient labeling, including adequate Contraindications, Warnings, and Precautions and description of the clinical events
- Additional warnings noting limitations of safety information (e.g., “The safety of the use of this device in [situation] has not been evaluated.”)
- Labeling the device for prescription use only
- Limitation to caregivers with certain qualifications or clinical training
- Limit to users with a minimum set of qualifications and/or training
- Physician/user training program
- Device tracking
- Other(s) [Click here to list other\(s\)](#)

1264 **Question 7: Can the risks be mitigated, so that Probable Benefits outweigh the Risks?**

1265 Indication

A B

- Yes –The probable benefits outweigh the risks such that, for this device, additional consideration of relevant factors would not change the determination.
- Undetermined – The probable benefits may not outweigh the risks, and further discussion and consideration of postmarket actions is appropriate → Move to Question 8

1266 **Comments on Assessment of Probable Benefit-Risk, considering risk mitigation strategies**

1267 **For the Proposed Indications for Use (Column A):**

1268 Click here to summarize the probable benefit(s) that have been demonstrated for the proposed Indications  
1269 for Use and your assessment of how Probable Benefit(s) compare to Risks. Include a description of how  
1270 available alternative modalities, including their probable benefits and risks, affect your assessment.

1271 Include a description of how uncertainty regarding Probable Benefit(s) and Risks affects your assessment.

1272 Include a description of how patient perspectives affected your assessment.

1273 **For the narrowed Indications for Use (modified indication and/or population) (Column B):**

1274 Click here to summarize the probable benefit(s) that have been demonstrated for the narrowed Indications  
1275 for Use and your assessment of how Probable Benefit(s) compare to Risks. Include a description of how  
1276 available alternative modalities, including their probable benefits and risks, affect your assessment.

1277 Include a description of how uncertainty regarding Probable Benefit(s) and Risks affects your assessment.

1278 Include a description of how patient perspectives affected your assessment.

1282 **8. Do the Probable Benefits outweigh the Risks considering the use of postmarket actions?**

1283 *Select appropriate postmarket action(s).*

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

1285 Indication

A B

- Collection of additional and/or confirmatory non-clinical performance data in the postmarket space
- Collection of additional and/or confirmatory clinical data in the postmarket space
- Other(s) [Click here to list other\(s\)](#)

1286  
1287 **Question 8: Do the Probable Benefits outweigh the Risks considering the use of postmarket**  
1288 **actions?**

1289 Indication

A B

- Yes – The probable benefits outweigh the risks.
- No – If you have determined that the probable benefits do not outweigh the risks, move to the right column in the table to assess the probable benefits and risks for a narrowed indication, or if the final column has been reached, and you have determined that the probable benefits do not outweigh the risks, do not approve the application.

1290  
1291 **Comments on the Assessment of Probable Benefit-Risk, considering postmarket actions**

1292 **For the Proposed Indications for Use (Column A):**

1293 Click here to summarize the probable benefits(s) that have been demonstrated for the proposed  
1294 Indications for Use and your assessment of how Probable Benefit(s) compare to Risks. Include a  
1295 description of how available alternative modalities, including their probable benefits and risks, affect your  
1296 assessment. Include a description of how uncertainty regarding Probable Benefit(s) and Risks affects  
1297 your assessment. Include a description of how patient perspectives affected your assessment.

1298  
1299 **For narrowed Indications for Use (modified indication and/or population) (Column B):**

1300 Click here to summarize the probable benefits(s) that have been demonstrated for the narrowed  
1301 Indications for Use and your assessment of how Probable Benefit(s) compare to Risks. Include a  
1302 description of how available alternative modalities, including their probable benefits and risks, affect your  
1303 assessment. Include a description of how uncertainty regarding Probable Benefit(s) and Risks affects  
1304 your assessment. Include a description of how patient perspectives affected your assessment.

1305

## Appendix C – Probable Benefit-Risk Assessment Tool

HDE Probable Benefit-Risk Assessment: Decision Support Tool HDE Questions Based on the totality of the data	A. Proposed Indications for Use	B. Potential Narrowing of Indications for Use <sup>1</sup>
		Is the narrowed indication clinically reasonable?
<b>Assessment of Probable Benefit</b>	Considering benefit in terms of <ul style="list-style-type: none"> <li>Type</li> <li>Magnitude</li> <li>Probability</li> <li>Duration of effects</li> </ul>	Considering benefit in terms of <ul style="list-style-type: none"> <li>Type</li> <li>Magnitude</li> <li>Probability</li> <li>Duration of effects</li> </ul>
1. Is there any evidence of clinical benefit?	<input type="checkbox"/> YES → Q2 <input type="checkbox"/> NO → move to B	<input type="checkbox"/> YES → Q2 <input type="checkbox"/> NO → Not approvable
2. What is the degree of uncertainty for the Benefits? <sup>2</sup>	<input type="checkbox"/> High <input type="checkbox"/> Medium <input type="checkbox"/> Low Continue to Q3	<input type="checkbox"/> High <input type="checkbox"/> Medium <input type="checkbox"/> Low Continue to Q3
<b>Assessment of Risk</b>	Considering risk in terms of <ul style="list-style-type: none"> <li>Severity, types, number and rates of harmful events</li> <li>Probability of a harmful event</li> <li>Duration of harmful events</li> <li>Risks from false-positive or false-negative results</li> </ul>	Considering risk in terms of <ul style="list-style-type: none"> <li>Severity, types, number and rates of harmful events</li> <li>Probability of a harmful event</li> <li>Duration of harmful events</li> <li>Risks from false-positive or false-negative results</li> </ul>
3. Are known/probable risks more than minimal?	<input type="checkbox"/> YES → Q4 <input type="checkbox"/> NO → Q4	<input type="checkbox"/> YES → Q4 <input type="checkbox"/> NO → Q4
4. What is the degree of uncertainty for the Risks?	<input type="checkbox"/> High <input type="checkbox"/> Med <input type="checkbox"/> Low Continue to Q5	<input type="checkbox"/> High <input type="checkbox"/> Med <input type="checkbox"/> Low Continue to Q5
<b>Assessment of Probable Benefit-Risk</b>		
5. Do the Probable Benefits outweigh the Risks? <sup>3</sup>	<input type="checkbox"/> YES → Worksheet complete <input type="checkbox"/> Undetermined → Q6	<input type="checkbox"/> YES → Worksheet complete <input type="checkbox"/> Undetermined → Q6
6. Do the Probable Benefits outweigh the Risks, taking into account additional relevant considerations?	<input type="checkbox"/> YES → Worksheet complete <input type="checkbox"/> Undetermined → Q7	<input type="checkbox"/> YES → Worksheet complete <input type="checkbox"/> Undetermined → Q7
7. Can the risks be mitigated, so that Probable Benefits outweigh the Risks?	<input type="checkbox"/> YES → Worksheet complete <input type="checkbox"/> Undetermined → Q8	<input type="checkbox"/> YES → Worksheet complete <input type="checkbox"/> Undetermined → Q8
8. Do the Probable Benefits outweigh the Risks considering the use of postmarket actions?	<input type="checkbox"/> YES → Worksheet complete <input type="checkbox"/> NO → move to B	<input type="checkbox"/> YES → Worksheet complete <input type="checkbox"/> NO → Not approvable

1306

<sup>1</sup> Instructions: The term “indications for use” describes the disease or condition that the device will diagnose, treat, prevent, cure, or mitigate, including a description of the patient population for which the device is intended. See 21 CFR 814.20(b)(3)(i) and 814.104(b)(4). Consider the probable benefits and risks for a modified population for the proposed use, a modified indication for the proposed population, or both a modified indication and modified population, which would translate into a ‘narrowing’ of the Indications for Use from what was originally proposed. Note that probable benefit and probable benefit-risk determinations for HDEs are different from those under PMAs. For more information, refer to Section VI of this guidance when it is finalized.

<sup>2</sup> Instructions: If the degree of uncertainty is low, then consider whether a different kind of marketing application would be appropriate. However, low uncertainty does not necessarily imply clinically meaningful benefit.

<sup>3</sup> Instructions: For an HDE, take into account the probable benefits and risks of currently available devices or alternative forms of treatment.